



CERTIFICATE OF CALIBRATION

Customer Name : Polo Labs Pvt. Ltd. Hoshiarpur

Model : Erba Chem 7x

Serial No. : N-1402122


Calibration Date : 11/04/2020

Calibration Due Date : 10/04/2021

This is to certify that we have calibrated the Semi Automated Biochemistry Clinical Chemistry Analyzer, Model: Erba Chem 7x bearing Sr. No. N-1402122 and on inspection of the instrument it is observed that the results are well within the range and instrument is working satisfactorily.

Thanking you,

For Transasia Bio-Medicals Ltd.


Rashpal Singh Rana
Sr. Regional Service Manager

M: +91-9316004670

Tel Fax No: +91-172-4643837

E mail: r.rana@transasia.co.in



UNMATCHED SERVICE
SINCE 1979...

CALIBRATION CERTIFICATE

1 NAME OF HOSPITAL : Polo Labs Pvt. Ltd, Hoshiarpur
2 MODEL : ERBA CHEM 5x
3 SR NO. : N-1402122
4 RANGE OF THE INSTRUMENT : OPTICAL DENSITY LINEAR FORM 0-2.5
5 DATE OF CALIBRATION : 11/04/2020
6 NEXT DUE DATE : 10/04/2021

TEST VOLTAGES / TEST TEMPERATURE

TEST VOLTAGE	ACTUAL VOLTAGE	OK/ NOT OK
LOGIC 5V	5.00V	OK
LAMP VOLTAGE (11.5V-12V)	11.75V	OK
SMPS V (17-19V)	18.30V	OK
FLOWCELL TEMPERATURE	25C,30C,37C, OFF	ALL OK
PUMP CALIBRATION		OK

TRANSASIA BIO-MEDICALS LTD.
Mahinder