

GST NO. : 08AAFCA6494E1ZJ
PAN NO. : AAFCA6494E

Ph. : 0141-2783950
Mob. : +91-95294-08805



AB HARSH ELECTRONIC (P) LTD.

1st Floor 50/41, Opp. Neerja Modi School, Rajat Path, Mansarovar, Jaipur-302020 • E-mail : info@abharsh.com

Calibration Detail

Lot.No. Plus0920 EXP. Date -05-10-2020

Instrument Model :- Mindray Bc-5130 Sr.No:- TR-75000049 Calibration Date 15-09-2020

Parameter	WBC	RBC	HGB	MCV	PLT
Target	8.67	4.53	13.7	91.0	252
1	7.61	4.15	12.9	89.5	233
2	7.75	4.13	13.2	89.4	232
3	7.79	4.08	13.0	88.9	229
4	7.51	4.06	12.5	88.9	231
5	7.48	4.08	12.5	88.7	232
6	7.69	4.14	12.5	88.5	234
7	7.31	4.10	12.5	88.8	227
8	7.46	4.15	12.5	89.1	239
9	7.52	4.11	12.5	89.1	236
10	7.51	4.15	12.6	88.6	233
Mean	7.51	4.12	12.6	88.9	233
CV%	1.05	0.8	1.1	0.4	1.5
NEW Factor(%)	115.43	110.03	109.03	102.36	108.10
Old Factor(%)	100.00	100.00	98.00	100.00	115.00

FOR A B HARSH ELECTRONICS (P) LTD.

