



# GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdl, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276470703 / 9028888728



CERTIFICATE NO. : CC-2957

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b> :- <b>OMSAI PATHOLOGY LABORATORY LATUR</b>	<b>Page No.</b> :- 1 of 1 <b>Certificate No.</b> :- GTS/201218/07-002 <b>Date of Received</b> :- 18.12.2020 <b>Date of Calibration</b> :- 18.12.2020 <b>Next Calibration Due On</b> :- 17.08.2021 <b>Issue Date</b> :- 31.12.2020 <b>Calibration method No.</b> :- MECH-WI-06 <b>ULR No</b> :- CC295720000011424P
<b>Ambient Temp. (°C)</b> :- 23.8 <b>Relative Humidity (%RH)</b> :- 49 <b>Barometric Pressure (mbar)</b> :- 943.9 <b>Location of calibration</b> :- In Lab <b>Condition of Item</b> :- Ok	

<b>2. Description of Item</b>			
<b>Name</b> :- Micropipette	<b>Range</b> :-	5 to 50 ul	
<b>Id No</b> :- OSPL/PIP/02	<b>Least Count</b> :-	1.0 ul	
<b>Make</b> :- Finnpiette	<b>Location</b> :-	Pathology	
<b>Type</b> :- Variable	<b>Sr No</b> :-	JW 01075	

<b>3.Details of Equipment used for calibration</b>				
<b>Name</b>	<b>Certificate No.</b>	<b>Certified By</b>	<b>ID/Sr. No.</b>	<b>Calibration Validity</b>
Weighing Balance	NI/GTS/030820/004	Nishitronics Instrumentation	GTS/WB-01	02.06.2021

<b>*Mechanical Calibration</b>				
<b>4.Calibration Results</b> :-				
<b>Calibration Points</b> ul	<b>Standard Reading</b> ul	<b>Set Value on UUC</b> ul	<b>Error in</b> ul	<b>Expanded</b> <b>Uncertainty in ± µl</b>
5	4.9930	5	0.0070	1.50
25	24.9835	25	0.0165	1.50
50	49.9260	50	0.0740	1.50

**Note:**  
 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution  
 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.  
 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.  
 4) Calibration point were selected as per customer specifications.  
 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".  
 6) 5ul not in NABL scope

Calibrated By  
  
 Calibration Engineer  
 D Ballal  
 RF-51/00



Approved By  
  
 Technical Manager  
 Swapnil Bhagawat

End of Certificate

VINO V-1 PRO  
2021.02.04 20:15

Email : globaltechnical007@gmail.com