



# GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276470703 / 9028888728



CERTIFICATE NO. : CC-2957

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b>	:-	Page No.	:- 1 of 1
<b>OMSAI PATHOLOGY LABORATORY</b>		Certificate No.	:- GTS/201218/07-001
<b>LATUR</b>		Date of Received	:- 18.12.2020
		Date of Calibration	:- 18.12.2020
		Next Calibration Due On	:- 17.06.2021
		Issue Date	:- 31.12.2020
Ambient Temp. (°C)	:- 23.8	Calibration method No.	:- MECH-WI-06
Relative Humidity (%RH)	:- 52	ULR No	:- CC295720000011423F
Barometric Pressure (mbar)	:- 942.2		
Location of calibration	:- In Lab		
Condition of Item	:- Ok		

<b>2. Description of Item</b>			
Name	:- Micropipette	Range	:- 200 ul
Id No	:- OSPL/PIP/01	Least Count	:- Nil
Make	:- Hemo Star	Location	:- Pathology
Type	:- Fixed	Sr No	:- NA 441759

<b>3.Details of Equipment used for calibration</b>				
Name	Certificate No.	Certified By	ID/Sr. No.	Calibration Validity
Weighing Balance	NI/GTS/030620/004	Nishitronics Instrumentation	GTS/WB-01	02.06.2021

### Mechanical Calibration

<b>4.Calibration Results</b>				
Calibration Points ul	Standard Reading ul	Set Value on UUC ul	Error in ul	Expanded Uncertainty in ± μl
200	199.8408	200	0.1592	1.50

**Note:**

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor  $k=2$ , which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".

Calibrated By  
D.B  
Calibration Engineer  
D Ballal  
RF-51/00



Approved By  
[Signature]  
Technical Manager  
Swapnil Bhagawat

End of Certificate

VIVO V11 PRO  
AI Dual Camera

2021.02.04 20:08