



# GLOBAL TECHNICAL SERVICES

Sec: No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276470703 / 9028888728



CERTIFICATE NO. : CC-2957

## CALIBRATION CERTIFICATE

<b>1.CUSTOMER</b>	-	Page No.	- 1 of 1
<b>OMSAI PATHOLOGY LABORATORY</b>		Certificate No.	- GTS/201218/07-004
<b>LATUR</b>		Date of Received	- 18.12.2020
		Date of Calibration	- 18.12.2020
		Next Calibration Due On	- 17.06.2021
		Issue Date	- 31.12.2020
Ambient Temp. (°C)	- 24.4	Calibration method No.	- MECH-WI-06
Relative Humidity (%RH)	- 50	ULR No	- CC295720000011426F
Barometric Pressure (mbar)	- 942.8		
Location of calibration	- In Lab		
Condition of Item	- Ok		

<b>2. Description of Item</b>			
Name	- Micropipette	Range	- 20 - 200 ul
Id No	- OSPL/PIP/04	Least Count	- 1 ul
Make	- ASTRA	Location	- Pathology
Type	- Variable	Sr No	- YN6A01614


<b>3.Details of Equipment used for calibration</b>				
Name	Certificate No.	Certified By	ID/Sr. No.	Calibration Validity
Weighing Balance	NI/GTS/030620/004	Nishitronics Instrumentation	GTS/WB-01	02.06.2021

### \*Mechanical Calibration

<b>4.Calibration Results</b>				
Calibration Points ul	Standard Reading ul	Set Value on UUC ul	Error in ul	Expanded Uncertainty in ± μl
20	19.9724	20	0.0276	1.50
100	99.9100	100	0.0900	1.50
200	199.8790	200	0.1210	1.50

**Note:**

- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".

Calibrated By  
  
 Calibration Engineer  
 D Ballal  
 RF-51/00



Approved By  
  
 Technical Manager  
 Swapnil Bhagawat

End of Certificate



Vivo V11Pro  
AI Dual Camera

Email : globaltechnical007@gmail.com

2021.02.04 20:14