



GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276470703 / 9028888728



CERTIFICATE NO. : CC-2957

CALIBRATION CERTIFICATE

1.CUSTOMER :-
OMSAI PATHOLOGY LABORATORY
LATUR

Page No. :- 1 of 1
Certificate No. :- GTS/201218/07-007
Date of Received :- 18.12.2020
Date of Calibration :- 18.12.2020
Next Calibration Due On :- 17.06.2021
Issue Date :- 31.12.2020
Calibration method No. :- TH-WI-02
ULR No :- CC295720000011429F

Ambient Temp. (°C) :- 28.5
Relative Humidity (%RH) :- 49
Location of calibration :- On Site
Condition of Item :- Ok

2. Description of Item

Name :- Digital Thermometer
Id No :- OSPL/DTHR/01
Make :- TPM - 10
Type :- Digital
Range :- -50 - 110 °C
Least Count :- 0.1 °C
Location :- Pathology

3.Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Precision Thermometer With Sensor	NI/2006/002/001	Nishitronics Instrumentation	GTS/PTH/01	01.06.2021

*Thermal Calibration

4.Calibration Results :-

Calibration Points °C	Corrected Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
-20	-20.05	-20	0.05	0.80
2	2.08	2	-0.08	0.80
5	5.14	5	-0.14	0.80
10	10.16	10	-0.16	0.80

Notes:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor $k=2$, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
- 6) UUC :- diameter : 3 mm & immersion depth : 40 mm

Calibrated By

D.B

Calibration Engineer

D Ballal

RE-5545



End of Certificate

Approved By

gnt

Technical Manager

Swapnil Bhagwat

Email : globaltechnical007@gmail.com

2021 02 04 20:19

VIVO V11Pro
AI Dual Camera

