



# GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276470703 / 9028888728



CERTIFICATE NO. : CC-2957

## CALIBRATION CERTIFICATE

1. CUSTOMER :-  
OMSAI PATHOLOGY LABORATORY  
LATUR

Page No. :- 1 of 1  
Certificate No. :- GTS/201218/07-008  
Date of Received :- 18.12.2020  
Date of Calibration :- 18.12.2020  
Next Calibration Due On :- 17.08.2021  
Issue Date :- 31.12.2020  
Calibration method No. :- TH-WI-02  
ULR No :- CC295720000011430F

Ambient Temp. (°C) :- 28.8  
Relative Humidity (%RH) :- 51  
Location of calibration :- On Site  
Condition of Item :- Ok

### 2. Description of Item

Name :- Digital Thermometer  
Id No :- OSP/DTHR/02  
Make :- TPM - 10  
Type :- Digital  
Range :- -50 - 110 °C  
Least Count :- 0.1 °C  
Location :- Pathology

### 3. Details of Equipment used for calibration

Name	Certificate No.	Certified By	ID No.	Calibration Validity
Precision Thermometer With Sensor	NI/2006/002/001	Nishitronics Instrumentation	GTS/PTH/01	01.06.2021

### \*Thermal Calibration

#### 4. Calibration Results :-

Calibration Points °C	Corrected Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
-20	-20.03	-20	0.03	0.80
2	2.06	2	-0.06	0.80
5	5.12	5	-0.12	0.80
10	10.18	10	-0.18	0.80

#### Note:

- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
- 6) UUC :- diameter : 3 mm & Immersion depth : 40 mm

Calibrated By

D-8

Calibration Engineer  
D Ballal



End of Certificate

Approved By

[Signature]

Technical Manager  
Swapnil Bhagwat

Email : globaltechnical007@gmail.com

Vivo V11 Pro  
ALDUAL Camera

2021/02/04 20:24