

CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/20-21/01072

Page: 1 of 2

Customer Name:	SREE METRO DIAGNOSTIC CENTRE	Customer Address:	16/41, KRISHNA ROAD, NEW PERUNGALATHUR, CHENNAI-63
Location:	LAB-1	Serial No:	1763325
Description:	WEIGHING SCALE	Tag/Asset No:	SMDC/LAB-1/WS/01
Make:	KRUPS	Date of Calibration:	30-09-2020
Model:	NA	Date of Due:	29-09-2021

Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1	Standard Weights 1kg	Sensortron	NA	033	12-08-2021
2.	Standard Weights 2kg	Sensortron	NA	034	12-08-2021
3.	Standard Weights 5kg	Sensortron	NA	305	31-03-2021

Environmental Conditions and Standard Operating Procedure Details:

Temperature	24°C	Calibration Done At	ONSITE
Humidity (40-70% RH)	55%	Calibration Procedure No	CP-BM010-2019/WM

REMARKS:

Performance Test Report : Attached

The above mentioned instruments has been calibrated using standard manufacturer recommended protocols, using equipments having traceability to National/International Standards. Partial reproduction of the certificates is not permitted.

Calibrated by:


M. BALAJI

(Calibration Engineer)

Authorized Signatory:


C. SHARMUGARAJ
(Technical Manager)

Result of Calibration: SBS/TR/20-21/01072


Page: 2 of 2

QUANTITATIVE TASKS						
S.No	Description	Units	Set Values	Measured Values	Limit/Tolerance	Result
1.	Weight	Kg	1	1.0	±0.5	PASS
2.	Weight	Kg	2	2.0	±0.5	PASS
3.	Weight	Kg	3	3.0	±0.5	PASS
4.	Weight	Kg	5	5.0	±0.5	PASS
5.	Weight	Kg	8	8.0	±0.5	PASS

Observation: The unit was found to delivery energy within allowed deviation limits.

CALIBRATION RESULT:

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: 

Date: 30-09-2020