

## VIEW LAB MONTHLY SUMMARY

Lab Name VSV CLINICAL LABORATORY

Lab No 9777

Month January

Year 2021

Constituent Group Chemistry II

[Details About Robust Analysis](#)[Detail About Monthly Summary](#)[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 19/01/2021

Date of Report Published : 05/02/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	695	119.49	9.92	11.85	134.2 mg/dl	1.24	0.90
2	UREA II	Urease UV / GLDH II	Transasia / Erba	509	31.68	12.76	4.04	30.4 mg/dl	-0.32	0.36
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	532	1.32	12.99	0.17	1.5 mg/dl	1.05	0.01
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	596	1.93	20.18	0.39	2.1 mg/dl	0.44	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	527	5.66	9.56	0.54	5.9 g/dl	0.44	0.05
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	527	3.38	9.61	0.32	3.7 g/dl	0.98	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	567	5.52	13.74	0.76	5.4 mg/dl	-0.16	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	637	111.91	12.83	14.35	110 mg/dl	-0.13	1.14
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	577	129.39	15.04	19.46	143.1 mg/dl	0.70	1.62

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>