

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCA8053H1ZR • DL No. MDU/4205/20B, MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. VSV Labs, No.45, Ponmeni Narayanan Street, S.S Colony, Madurai - 625 016	Certificate No:	AMS/LAB/2888496
		Issuing Date:	26/11/2020
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Centrifuge	Id. No:	VSV/LAB/CEN/001
Make:	REMI	Department:	Laboratory
Model:	C-854/6	Calibrated On:	25/11/2020
Sl. No:	11240	Calibration	
Type:	Manual	Due Suggested:	24/11/2021
Range & Least Count:	0 - 3500 / 0.65A	DUC Condition on Receipt:	OK

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Digital Tachometer	Lutron / DT-2234C	FSMS/CL/05/18-08/071-01	29/07/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.No.	DUC Reading positions	Standard Reading (RPM)	Correction (RPM)
1	Position 1	1843	0
2	Position 2	2179	0
3	Position 3	3050	0
4	Position 4	3127	0
5	Position 5	3243	0

Allowable Tolerance: Set by Customer

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. S. S.

© aceagm@gmail.com



Approved by:

(Technical Manager)

B. S. S.

©www.acemedical.in

☎ 93455 56140, 93455 56148