Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No: 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone: 0452 - 2561691 Mobile: +91 94434 74610



GSTIN. 33AAHCA8053H1ZR

DL No. MDU/4205/20B,MDU/3971/21B

CIN: U33110TN2009PTC072140



CALIBRATION CERTIFICATE

Customer	M/S. VSV Labs,	Certificate No:	AMS/LAB/2888503
Name &	No.45, Ponmeni Narayanan Street,	Issuing Date:	26/11/2020
Address:	S.S Colony, Madurai – 625 016	Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micro Pipette	Id. No:	VSV/LAB/FMP/004
Make:	Diagnopette	Department:	Laboratory
Model:	Fixed Pipette	Calibrated On:	25/11/2020
Sl. No:		Calibration	
Type:	Fixed	Due Suggested:	24/11/2021
Range &	500μ1	DUC Condition	
Least Count:		on Receipt:	OK

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	LM-2008863	08/08/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.	DUC Reading	Standard Reading (µI)		Correction
No.	(µl)	Trial 1	Trial 2	(μl)
1	500	500	500	0

Allowable Tolerance: ±1%

Note:

- 1. Calibration Points were selected as per customer request.
- 2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
- 3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)





Approved by:

(Technical Manager)