

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Sobhi Diagnostic Centre, Main Road, North Bazar, Anjugramam - 629 401, Kanyakumari Dist.	Certificate No:	AMS/LAB/2889156
		Issuing Date:	27/01/2021
		Page No:	1 of 2

Details of Device under Calibration (DUC):

DUC:	Bio Chemistry Analyser	Id. No:	SDCA/LAB /BCA/001
Make:	ROBONIK	Department:	Laboratory
Model:	Prietest Compact	Calibrated On:	26/01/2021
Sl. No:	PC 0240115 RBK	Calibration	
Type:	Semi Auto	Due Suggested:	25/01/2022
Range & Least Count:	---	DUC Condition on Receipt:	OK

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/5507/2021	18/01/2022
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	275	mA	500	Normal
2	Frequency	50.09	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	201.53	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	210.8	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	9.9	Volts	< 7	Abnormal

*Electrical Safety: Please check the attached log sheet

SL.No	Parameter	Measured	Standard	Correction	Unit
1	Lamp Voltage	11.48	12.00	-0.52	Volts
2	Detector Sensitivity	-0.00	0.0	0.0	% Absorbance
3	Sipper Volume	341.25	400	-58.75	ml
4	Peristaltic Pump pressure	-12.80	-15.0	-2.20	cmH2O

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

B. S. S.



Approved by:
(Technical Manager)

B. S. S.

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company


No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. Sobhi Diagnostic Centre, Main Road, North Bazar, Anjugramam – 629 401, Kanyakumari Dist.	Certificate No:	AMS/LAB/2889162
		Issuing Date:	27/01/2021
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Refrigerator – Reagent	Id. No:	SDCA/LAB/REFR/001
Make:	LG	Department:	Laboratory
Model:	GL B181RCOM/2016	Calibrated On:	26/01/2021
Sl. No:	610NRAL078520	Calibration Due Suggested:	25/01/2022
Type:	Reagents	DUC Condition on Receipt:	OK
Range & Least Count:	0 – 8°C / LC 1°C		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	QC/AMS/DTI/002	22/02/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

Test Results:

SL.No.	Place of Reading	Standard Reading (°C)	Correction (°C)
1	Chiller	-1.26	0
2	Chamber	2.67	0

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

B. S. S.



☉ aceagm@gmail.com

☉ www.acemedical.in



Approved by:
(Technical Manager)

B. S. S.

☎ 93455 56140, 93455 56148

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610




GSTIN. 33AAHCA8053H1ZR

DL No. MDU/4205/20B,MDU/3971/21B

CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. Sobhi Diagnostic Centre, Main Road, North Bazar, Anjugramam – 629 401, Kanyakumari Dist.	Certificate No:	AMS/LAB/2889157
		Issuing Date:	27/01/2021
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Centrifuge	Id. No:	SDCA/LAB/CEN/001
Make:	Remi	Department:	Laboratory
Model:	C852/4	Calibrated On:	26/01/2021
Sl. No:	BFLC – 399	Calibration Due Suggested:	25/01/2022
Type:	Manual	DUC Condition on Receipt:	OK
Range & Least Count:	0 – 3500RPM / LC - 1		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Digital Tachometer	Lutron / DT-2234C	FSMS/CL/05/19-07/203-01	29/07/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.No.	DUC Reading (Position)	Standard Reading (RPM)	Correction (RPM)
1	Position 1	1176	0
2	Position 2	1752	0
3	Position 3	2128	0
4	Position 4	2794	0
5	Position 5	3157	0

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. S. S.



Approved by:

(Technical Manager)

B. S. S.

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company


No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. Sobhi Diagnostic Centre, Main Road, North Bazar, Anjugramam – 629 401, Kanyakumari Dist.	Certificate No:	AMS/HOSP/2889159
		Issuing Date:	27/01/2021
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micro Pipette	Id. No:	SDCA/LAB/FMP/001
Make:	Lab Auto Pipette	Department:	Laboratory
Model:	Fixed	Calibrated On:	26/01/2021
Sl. No:	---	Calibration Due Suggested:	25/01/2022
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	10µl		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	LM-2008863	08/08/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

Test Results:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	10	10	10	0

Allowable Tolerance: ±1%

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

B. Srinivas



www.acemedical.in



Approved by:
(Technical Manager)

B. Srinivas

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. Sobhi Diagnostic Centre, 18/4, Perumal Coil street, Vallioor - 627 117, Tirunelveli Dist.	Certificate No:	AMS/HOSP/2889160
		Issuing Date:	27/01/2021
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micro Pipette	Id. No:	SDCA/LAB/VMP/001
Make:	Labopette	Department:	Laboratory
Model:	Variable	Calibrated On:	26/01/2021
Sl. No:	---	Calibration Due Suggested:	25/01/2022
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	10 to 100µl		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	LM-2008863	08/08/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	10	10	10	0
2	20	20	20	0
3	50	50	50	0
4	100	100	100	0

Allowable Tolerance: ±1%

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. S. S.



www.acemedical.in



Approved by:

(Technical Manager)

B. S. S.

aceagm@gmail.com

93455 56140, 93455 56148

Scanned with CamScanner

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR

DL No. MDU/4205/20B, MDU/3971/21B

CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. Sobhi Diagnostic Centre, Main Road, North Bazar, Anjugramam – 629 401, Kanyakumari Dist.	Certificate No:	AMS/HOSP/2889161
		Issuing Date:	27/01/2021
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micro Pipette	Id. No:	SDCA/LAB/VMP/002
Make:	Lab Auto Pipette	Department:	Laboratory
Model:	Variable	Calibrated On:	26/01/2021
Sl. No:	---	Calibration Due Suggested:	25/01/2022
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	100 to 1000µl		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	LM-2008863	08/08/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	100	100	100	0
2	200	200	200	0
3	500	500	500	0
4	1000	1000	1000	0

Allowable Tolerance: ±1%

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. S. S.



Approved by:
(Technical Manager)

B. S. S.