

SERVICE REPORT

Date: 25/10/2021

LMD#

1317

Complaint Date: 25/10/2021Job Completion Date: 25/10/2021

Type of Call	DEFECT DEVICE <input type="checkbox"/>	MAINTENANCE <input checked="" type="checkbox"/>	INSTALLATION <input type="checkbox"/>
	MAINTENANCE <input checked="" type="checkbox"/>	CALL <input type="checkbox"/>	REPAIR WITH SERVICE ON CALL <input checked="" type="checkbox"/>

Hospital - Laboratory: 3 Best Biomedical Systems Private Limited
 (Name & Address):
85, 8th main paragraph Area Pragas Bangalore - 56

Contact Person: Dr. Manjunath S
 Tel:
 Fax:
 Mob: 9962580011
 Email:

Instrument: Controlux, Model: Pony
Refugee, serial No: 10110

Priority of Complaint:
 (Indicate the priority of the complaint in a separate sheet with the reason):
calibration done
 P.A. Voltage: 00 N.C. Voltage: 00 O.P. Voltage: 00

Equipment Observation: <u>Controlux meter</u> <u>Power Refugee meter</u> <u>calibrated with the measuring</u> <u>instrument</u>	Remarks: <u>Controlux, 10110 power</u> <u>and Refugee meter calibration</u> <u>done</u>
---	--

Previous service: YES NO
 Type of Stabilizer: VPS

Parts Replaced: <u>no</u>	Warranty Details: <u>no</u>
------------------------------	--------------------------------

Engineer: Jyoti K Customer Signature & Seal:

Terms & Conditions

- The warranty will not be extended to & the change in specification of scope will not included in the service charges.
- Company is not responsible for any physical damage caused to the instrument by the negligence of user. The damage caused due to natural calamities.
- Our work on the year equipment will 10% charge for spare parts necessary condition.
- All specifications under Bangalore for service.