

SPARSHA DIAGNOSTICS (430)

Outlier And Analyte Summary Report

Result Details For Cycle No 200312 and Sample No 04

Report Date : 14/01/2021

Instrument : Toshiba TBA-25FR

| Analyte | StandardUnit | Result Value | Mean | Z-Score | RMZ |
|-------------------------|--------------|--------------|------|---------|-------|
| @ Glucose | mg/dl | 70.90 | -- | -0.17 | -0.27 |
| @ Urea | mg/dl | 20.90 | -- | 0.00 | 0.53 |
| @ Creatinine | mg/dl | 0.80 | -- | 1.82 | 1.82 |
| @ Phosphorous Inorganic | mg/dl | 3.60 | -- | -0.30 | -0.51 |
| @ Cholesterol | mg/dl | 140.00 | -- | -1.02 | -1.13 |
| @ HDL | mg/dl | 32.50 | -- | -0.06 | -0.07 |
| @ LDL | mg/dl | 98.51 | -- | -0.24 | -0.18 |

Legend @ : z score \leq 2.0 - Acceptable
 | : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

| | |
|--------------------------|----------|
| Total Parameters | 7 |
| Not Evaluated Parameters | 0 |
| Evaluated Parameters | 7 |
| Outlier Parameters (X) | 0 |
| EQAS Score Biochemistry | 100.00 % |

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

End of Report

SPARSHA DIAGNOSTICS (430)

Outlier And Analyte Summary Report

Result Details For Cycle No 200312 and Sample No 01

Report Date : 14/01/2021

Instrument : CKK LYE - ARK DIAGNOSTICS

| Analyte | StandardUnit | Result Value | Mean | Z-Score | RMZ |
|-------------|--------------|--------------|------|---------|-------|
| @ Chloride | mmol/L | 106.00 | -- | -0.71 | -0.77 |
| X Sodium | mmol/L | 178.90 | -- | 9.51 | 8.29 |
| X Potassium | mmol/L | 7.63 | -- | 11.32 | 11.78 |

Instrument : Toshiba TBA-120FR

| Analyte | StandardUnit | Result Value | Mean | Z-Score | RMZ |
|-------------|--------------|--------------|------|---------|-------|
| X Magnesium | mg/dl | 1.30 | -- | -3.34 | -2.90 |

Instrument : Toshiba TBA-25FR

| Analyte | StandardUnit | Result Value | Mean | Z-Score | RMZ |
|--------------------|--------------|--------------|------|---------|-------|
| @ Bilirubin Total | mg/dl | 0.26 | -- | 0.51 | 0.55 |
| @ Bilirubin Direct | mg/dl | 0.09 | -- | -0.11 | -0.14 |
| @ Total Protein | g/dL | 6.70 | -- | 1.64 | 1.24 |
| @ Albumin | g/dL | 3.40 | -- | -1.12 | -0.53 |
| @ Uric Acid | mg/dl | 3.74 | -- | -1.65 | -1.18 |
| @ Calcium | mg/dl | 9.00 | -- | -0.21 | -0.19 |
| @ Triglycerides | mg/dl | 122.70 | -- | -1.32 | -1.05 |
| ! ALT | U/L | 27.70 | -- | 2.72 | 1.35 |
| ! AST | U/L | 39.60 | -- | 2.17 | 1.37 |
| X ALP | U/L | 178.61 | -- | 8.52 | 3.73 |
| @ CK | U/L | 120.20 | -- | 0.91 | 0.86 |
| @ GGT | U/L | 24.40 | -- | -0.75 | -0.15 |
| @ LDH | U/L | 342.00 | -- | 1.28 | 1.30 |
| @ Amylase | U/L | 48.20 | -- | -1.90 | -2.01 |

Legend @ : z score \leq 2.0 - Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ! : Delayed Result Entry
 * : Not considered for evaluation.

| | |
|--------------------------|---------|
| Total Parameters | 18 |
| Not Evaluated Parameters | 0 |
| Evaluated Parameters | 18 |
| Outlier Parameters (X) | 4 |
| EQAS Score Biochemistry | 77.78 % |

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinoor Mall, Kirof Road, Kurla (W),
 Mumbai - 400070

SPARSHA DIAGNOSTICS (430)

Outlier And Analyte Summary Report

Result Details For Cycle No 200312 and Sample No 03

Report Date : 14/01/2021

Instrument : Toshiba TBA-25FR

| Analyte | StandardUnit | Result Value | Mean | Z-Score | RMZ |
|-------------------------|--------------|--------------|------|---------|-------|
| @ Glucose | mg/dl | 73.70 | -- | -0.37 | -0.27 |
| @ Urea | mg/dl | 23.30 | -- | 1.06 | 0.53 |
| @ Creatinine | mg/dl | 0.79 | -- | 1.81 | 1.82 |
| @ Phosphorous Inorganic | mg/dl | 3.30 | -- | -0.71 | -0.51 |
| @ Cholesterol | mg/dl | 139.00 | -- | -1.24 | -1.13 |
| @ HDL | mg/dl | 34.40 | -- | -0.07 | -0.07 |
| @ LDL | mg/dl | 97.86 | -- | -0.11 | -0.13 |

Legend @ : z score \leq 2.0 - Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

| | |
|--------------------------|----------|
| Total Parameters | 7 |
| Not Evaluated Parameters | 0 |
| Evaluated Parameters | 7 |
| Outlier Parameters (X) | 0 |
| EQAS Score Biochemistry | 100.00 % |

Dr Puneet Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kuria (W),
 Mumbai - 400070

End of Report

SPARSHA DIAGNOSTICS (430)

Outlier And Analyte Summary Report

Result Details For Cycle No 200312 and Sample No 02

Report Date : 14/01/2021

Instrument : CCK LYE - ARK DIAGNOSTICS

| Analyte | StandardUnit | Result Value | Mean | Z-Score | RMZ |
|-------------|--------------|--------------|------|---------|-------|
| @ Chloride | mmol/L | 106.30 | -- | -0.83 | -0.77 |
| X Sodium | mmol/L | 181.00 | -- | 7.06 | 8.29 |
| X Potassium | mmol/L | 7.94 | -- | 12.24 | 11.78 |

Instrument : Toshiba TBA-120FR

| Analyte | StandardUnit | Result Value | Mean | Z-Score | RMZ |
|-----------|--------------|--------------|------|---------|-------|
| Magnesium | mg/dl | 1.40 | -- | -2.46 | -2.90 |

Instrument : Toshiba TBA-25FR

| Analyte | StandardUnit | Result Value | Mean | Z-Score | RMZ |
|--------------------|--------------|--------------|------|---------|-------|
| @ Bilirubin Total | mg/dl | 0.25 | -- | 0.58 | 0.55 |
| @ Bilirubin Direct | mg/dl | 0.08 | -- | -0.17 | -0.14 |
| @ Total Protein | g/dL | 6.59 | -- | 0.84 | 1.24 |
| @ Albumin | g/dL | 3.70 | -- | 0.07 | -0.53 |
| @ Uric Acid | mg/dl | 3.75 | -- | -0.71 | -1.18 |
| @ Calcium | mg/dl | 9.10 | -- | -0.16 | -0.19 |
| @ Triglycerides | mg/dl | 113.50 | -- | -0.78 | -1.03 |
| @ ALT | U/L | 15.40 | -- | -0.02 | 1.35 |
| @ AST | U/L | 21.80 | -- | 0.57 | 1.37 |
| X ALP | U/L | 177.59 | -- | 8.94 | 8.73 |
| @ CK | U/L | 101.70 | -- | 0.81 | 0.86 |
| @ GGT | U/L | 21.10 | -- | 0.46 | -0.15 |
| @ LDH | U/L | 333.00 | -- | 1.31 | 1.30 |
| Amylase | U/L | 43.30 | -- | -2.12 | -2.01 |

Legend @ : z score \leq 2.0 - Acceptable
 : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

| | |
|--------------------------|---------|
| Total Parameters | 18 |
| Not Evaluated Parameters | 0 |
| Evaluated Parameters | 18 |
| Outlier Parameters (X) | 3 |
| EQAS Score Biochemistry | 83.33 % |

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

End of Report