



CALIBRATION CERTIFICATE

1. CUSTOMER BANSAL HOSPITAL SHAHPURA, BHOPAL-462016. Ambient Temp :- (25 ± 4)°C RH :- < 70% Location of calibration :- ON SITE Condition of Item :- OK *	Page No :- 1 of 1 Date of Issue :- 9/9/2020 Service Request No. :- -- Certificate No. :- REF-18 Date of Calibration :- 9/9/2020 Next Calibration Due On :- 8/9/2021 Calibration method No. :- ACS/CM-101 Discipline :- Thermal Calibration Parameter :- Temp
---	--


2. Description of Item	
Name :- TEMPERATURE INDICATOR Sr.No :- FKG120180900007 Make :- CELFROST Model :- N.A Type :- DIGITAL Asset no :- N.A	Range :- 2 TO 8 °C Resolution :- 1 °C Specified Accuracy :- N.A Location :- REFRIGERATOR Department :- 8TH FLOOR

3. Detail of Equipment used for calibration	
Name :- Universal Calibrator Make/I.D No. :- Massibus/ACS-EQP-06 Certificate No. :- CC/ECL/1442/19-20 Certified By :- IDEMI Calibration Validity :- 13-Nov-20	PT-100 Sensor Tempsens/ACS-EQP-05 CC/TRL/0607/19-20 IDEMI 17-Nov-20

4. Calibration Results :-				
Calibration Points °C	Converted Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	1.6	2	0.4	0.50
4	3.6	4	0.4	0.50
6	5.6	6	0.4	0.50
7	6.6	7	0.4	0.50
8	7.6	8	0.4	0.50

- Note:**
- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to coverage probability of approximately 95% for normal distribution.
 - 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
 - 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
 - 4) Calibration point were selected as per customer specifications and Calibration report traceable to National or International Standard.
 - 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Autocal Systems, Pune"

Calibrated By: *Anjali*
 Calibration Engineer



Approved By: *Asagde*
 Technical Manager
 (Mr. Ankush Bagde)

Freeze thermometer



AUTOCAL SYSTEMS

Calibration of Temperature, Length, Mass, Volume, Weights, Mapping and Yoking, Force Testing Services, Calibration of Environmental Test Chambers, Humidity Calibrators, OT Validation, AHU Calibration, Calibration of Test Cells, Creep



NABL ACCREDITED CALIBRATION LAB WITH ISO/IEC 17025

CALIBRATION CERTIFICATE

1. CUSTOMER BANSAL HOSPITAL SHAHPURA, BHOPAL-462016. Ambient Temp. :- (25 ± 4)°C RH :- < 70% Location of calibration :- ON SITE Condition of Item :- OK	Page No. :- 1 of 1 Date of Issue :- 9/9/2020 Service Request No. :- Certificate No. :- REF-22 Date of Calibration :- 9/9/2020 Next Calibration Due On :- 8/9/2021 Calibration method No. :- ACS/CM-101 Discipline :- Thermal Calibration Parameter :- Temp
--	--

2. Description of Item	
Name :- TEMPERATURE INDICATOR	Range :- 2 TO 8 °C
Sr.No :- 120913261592406981	Resolution :- 1 °C
Make :- ELECTROLUX	Specified Accuracy :- N.A
Model :- N.A	Location :- REFRIGERATOR
Type :- DIGITAL	Department :- 8TH FLOOR
Asset no :- N.A	

3.Detail of Equipment used for calibration	
Name :- Universal Calibrator	PT-100 Sensor
Make/I.D No. :- Massibus/ACS-EQP-06	Tempsans/ACS-EQP-05
Certificate No. :- CC/ECL/1442/19-20	CC/TRL/0607/19-20
Certified By :- IDEMI	IDEMI
Calibration Validity :- 13-Nov-20	17-Nov-20

4.Calibration Results :-				
Calibration Points °C	Converted Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	1.6	2	0.4	0.50
4	3.6	4	0.4	0.50
6	5.6	6	0.4	0.50
7	6.6	7	0.4	0.50
8	7.6	8	0.4	0.50

Note:

- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to coverage probability of approximately 95% for normal distribution.
- 2) This certificate refers only to the particular item submitted for calibration.UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications and Calibration report traceable to National or International Standard.
- 5) This certificate shall not be reproduced,except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Autocal Systems,Pune"

Calibrated By
Govali
Calibration Engineer



Approved By
Mr. Ankush Bagde
Technical Manager
(Mr.Ankush Bagde)

Three Clauses



NABL ACCREDITED CALIBRATION LAB WITH ISO/IEC 17025

CALIBRATION CERTIFICATE

1. CUSTOMER
 BANSAL HOSPITAL
 SHAHPURA,
 BHOPAL-462016.

Ambient Temp. :- (25 ± 4)°C
 RH :- < 70%
 Location of calibration :- ON SITE
 Condition of Item :- OK

Page No. :- 1 of 1
 Date of Issue :- 7/9/2020
 Service Request No. :-
 Certificate No. :- REF-20
 Date of Calibration :- 7/9/2020
 Next Calibration Due On :- 6/9/2021
 Calibration method No. :- ACS/CM-101
 Discipline :- Thermal Calibration
 Parameter :- Temp

2. Description of Item

Name :- TEMPERATURE INDICATOR
 Sr.No :- R5824APAD900294R
 Make :- CELFROST
 Model :- N.A
 Type :- DIGITAL
 Asset no :- N.A

Range :- 2 TO 8 °C
 Resolution :- 1 °C
 Specified Accuracy :- N.A
 Location :- REFRIGERATOR
 Department :- 8TH FLOOR

3. Detail of Equipment used for calibration

Name :- Universal Calibrator
 Make/I.D No. :- Massibus/ACS-EQP-06
 Certificate No. :- CC/ECL/1442/19-20
 Certified By :- iDEMi
 Calibration Validity :- 13-Nov-20

PT-100 Sensor
 Tempsens/ACS-EQP-05
 CC/TRL/0607/19-20
 iDEMi
 17-Nov-20


4. Calibration Results

Calibration Points °C	Converted Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	1.3	2	0.7	0.50
4	3.3	4	0.7	0.50
6	5.3	6	0.7	0.50
7	6.4	7	0.6	0.50
8	7.4	8	0.6	0.50


Note:

- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to coverage probability of approximately 95% for normal distribution.
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications and Calibration report traceable to National or International Standard.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Autocal Systems, Pune"

Calibrated By: *Arwuli*
 Calibration Engineer



Approved By: *Mr. Ankush Bagde*
 Technical Manager
 (Mr. Ankush Bagde)



CALIBRATION CERTIFICATE

1. CUSTOMER
BANSAL HOSPITAL
 SHAHPURA,
 BHOPAL-462016
 Ambient Temp :- (25 ± 4)°C
 RH :- < 70%
 Location of calibration :- ON SITE
 Condition of Item :- OK

Page No - 1 of 1
 Date of Issue - 9/9/2020
 Service Request No. - --
 Certificate No. - REF-18
 Date of Calibration - 9/9/2020
 Next Calibration Due On - 8/9/2021
 Calibration method No. - ACS/CM-101
 Discipline - Thermal Calibration
 Parameter - Temp

2. Description of Item

Name :- TEMPERATURE INDICATOR Range :- 2 TO 8 °C
 Sr.No :- FKG120180900007 Resolution :- 1 °C
 Make :- CELFROST Specified Accuracy :- N.A
 Model :- N.A Location :- REFRIGERATOR
 Type :- DIGITAL Department :- 8TH FLOOR
 Asset no :- N.A

3. Detail of Equipment used for calibration

Name :- Universal Calibrator PT-100 Sensor
 Make/I.D No. :- Massibus/ACS-EQP-06 Tempsens/ACS-EQP-05
 Certificate No. :- CC/ECL/1442/19-20 CC/TRL/0607/19-20
 Certified By :- IDEMI IDEMI
 Calibration Validity :- 13-Nov-20 17-Nov-20


4. Calibration Results :-

Calibration Points °C	Converted Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	1.6	2	0.4	0.50
4	3.6	4	0.4	0.50
6	5.6	6	0.4	0.50
7	6.6	7	0.4	0.50
8	7.6	8	0.4	0.50

Note:

- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to coverage probability of approximately 95% for normal distribution.
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications and Calibration report traceable to National or International Standard.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Autocal Systems, Pune"

Calibrated By: *Anjali*
 Calibration Engineer



Approved By: *Abagde*
 Technical Manager
 (Mr. Ankush Bagde)

Freeze the 0.00 meter



AUTOCAL SYSTEMS

Calibration of Biomedical, QA of Radiology, Thermal, Mechanical, Electrotechnical, Mass, Volume, Weights, Mapping and Validation Testing Services of Laminar Air Flow, Biosafety Cabinet, OT Validation, AHU and Calibration Facility in Lab / Onsite.



NABL ACCREDITED CALIBRATION LAB WITH ISO/IEC 17025

CALIBRATION CERTIFICATE

1. CUSTOMER
 BANSAL HOSPITAL
 SHAHPURA,
 BHOPAL-462016
 Ambient Temp :- (25 ± 4)°C
 RH :- < 70%
 Location of calibration :- ON SITE
 Condition of Item :- OK

Page No :- 1 of 1
 Date of Issue :- 9/9/2020
 Service Request No :-
 Certificate No :- REF-21
 Date of Calibration :- 9/9/2020
 Next Calibration Due On :- 8/9/2021
 Calibration method No :- ACS/CM-101
 Discipline :- Thermal Calibration
 Parameter :- Temp

2. Description of Item

Name :- TEMPERATURE INDICATOR Range :- 2 TO 8 °C
 Sr.No :- 131253917 Resolution :- 1 °C
 Make :- CELFROST Specified Accuracy :- N.A
 Model :- N.A Location :- REFRIGERATOR
 Type :- DIGITAL Department :- 7TH FLOOR
 Asset no :- N.A

3. Detail of Equipment used for calibration

Name :- Universal Calibrator Temp. indi with RTD sensor
 Make/I.D No. :- Massibus/ACS-EQP-06 Tempsens/ACS-EQP-05
 Certificate No. :- CC/ECL/1442/19-20 CC/TRL/0607/19-20
 Certified By :- IDEMI IDEMI
 Calibration Validity :- 13-Nov-20 17-Nov-20

4. Calibration Results

Calibration Points	Converted Standard Reading	UUC Reading	Error in	Expanded
°C	°C	°C	°C	Uncertainty in ± °C
2	1.5	2	0.5	0.50
4	3.6	4	0.4	0.50
6	5.5	6	0.5	0.50
7	6.4	7	0.6	0.50
8	7.5	8	0.5	0.50


Note:

- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to coverage probability of approximately 95% for normal distribution.
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications and Calibration report traceable to National or International Standard.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Autocal Systems, Pune"

Calibrated By:

Anjali

Calibration Engineer



Approved By:

Mr. Ankush Bagde

Technical Manager
 (Mr. Ankush Bagde)

83/526, Near D. Y. Patil Engg. College, Sant Tukaram Nagar, Pimpri, Pune - 411018.

☎ 09822136161 / 09975706701 / 07722092715 ✉ autocalsystemsmed@gmail.com / autocalsystems@gmail.com

www.autocalsystems.com



AUTOCAL SYSTEMS

Calibration of Biomedical, QA of Radiology, Thermal, Mechanical, Electrotechnical, Mass, Volume, Weights, Mapping and Validation Testing Services of Laminar Air Flow, Biosafety Cabinet, OT Validation, AHU and Calibration Facility in Lab / Onsite.



NABL ACCREDITED CALIBRATION LAB WITH ISO/IEC 17025

CALIBRATION CERTIFICATE

1. CUSTOMER

BANSAL HOSPITAL
SHAHPURA,
BHOPAL-462016
Ambient Temp. :- (25 ± 4)°C
RH :- < 70%
Location of calibration :- ON SITE
Condition of Item :- OK

Page No. :- 1 of 1
Date of Issue :- 7/9/2020
Service Request No. :-
Certificate No. :- REF-14
Date of Calibration :- 7/9/2020
Next Calibration Due On :- 6/9/2021
Calibration method No. :- ACS/CM-101
Discipline :- Thermal Calibration
Parameter :- Temp

2. Description of Item

Name :- TEMPERATURE INDICATOR WITH SENSOR Range :- 2 TO 8 °C
Sr.No :- 53024DBDA00160K Resolution :- 1 °C
Make :- SAMSUNG Specified Accuracy :- N.A
Model :- N.A Location :- REFRIGERATOR
Type :- DIGITAL Department :- IPD PHARMACY
Asset no :- N.A

3. Detail of Equipment used for calibration

Name :- Universal Calibrator PT-100 Sensor
Make/ID No. :- Massibus/ACS-EQP-06 Tempsens/ACS-EQP-05
Certificate No. :- CC/ECL/1442/19-20 CC/TRL/0607/19-20
Certified By :- IDEMI IDEMI
Calibration Validity :- 13-Nov-20 17-Nov-20

4. Calibration Results


Calibration Points °C	Converted Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	1.7	2	0.3	0.50
4	3.7	4	0.3	0.50
6	5.7	6	0.3	0.50
7	6.7	7	0.3	0.50
8	7.7	8	0.3	0.50

Note:

- The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to coverage probability of approximately 95% for normal distribution.
- This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- Calibration point were selected as per customer specifications and Calibration report traceable to National or International Standard.
- This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Autocal Systems, Pune"

Calibrated By:

Calibration Engineer



Approved By:

Technical Manager
(Mr. Ankush Bagde)

83/526, Near D. Y. Patil Engg. College, Sant Tukaram Nagar, Pimpri, Pune - 411018.

09822136161 / 09975706701 / 07722092715 autocalsystemsmmed@gmail.com / autocalsystems@gmail.com



AUTOCAL SYSTEMS

Calibration of Biomedical, QA of Radiology, Thermal, Mechanical, Electrotechnical, Mass, Volume, Weights, Mapping and Validation Testing Services of Laminar Air Flow, Biosafety Cabinet, OT Validation, AHU and Calibration Facility in Lab / Onsite.



NABL ACCREDITED CALIBRATION LAB WITH ISO/IEC 17025

CALIBRATION CERTIFICATE

1. CUSTOMER :-
BANSAL HOSPITAL
SHAHPURA,
BHOPAL-462016
Ambient Temp :- (25 ± 4)°C
RH :- < 70%
Location of calibration :- ON SITE
Condition of Item :- OK

Page No. :- 1 of 1
Date of Issue :- 7/9/2020
Service Request No. :-
Certificate No. :- REF-01
Date of Calibration :- 7/9/2020
Next Calibration Due On :- 8/9/2021
Calibration method No. :- ACS/CM-101
Discipline :- Thermal Calibration
Parameter :- Temp

2. Description of Item

Name :- TEMPERATURE INDICATOR WITH SENSOR Range :- 2 TO 8 °C
S.No :- R242233 Resolution :- 1 °C
Make :- SAMSUNG Specified Accuracy :- N.A
Model :- N.A Location :- REFRIGERATOR
Type :- DIGITAL Department :- CENTRAL PHARMACY
Asset no :- N.A

3. Detail of Equipment used for calibration

Name :- Universal Calibrator Temp. indi with RTD sensor
Make/I.D No. :- Massibus/ACS-EQP-06 Tempsens/ACS-EQP-05
Certificate No. :- CC/ECL/1442/19-20 CC/TRL/0607/19-20
Certified By :- IDEMI IDEMI
Calibration Validity :- 13-Nov-20 17-Nov-20


4. Calibration Results :-

Calibration Points	Converted Standard Reading	UUC Reading	Error in	Expanded
°C	°C	°C	°C	Uncertainty in ± °C
2	1.6	2	0.4	0.50
4	3.6	4	0.4	0.50
6	5.6	6	0.4	0.50
7	6.7	7	0.3	0.50
8	7.7	8	0.3	0.50

Note:

- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to coverage probability of approximately 95% for normal distribution.
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications and Calibration report traceable to National or International Standard.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Autocal Systems, Pune"

Calibrated By:
Calibration Engineer



Approved By:
Technical Manager
(Mr. Ankush Bagde)

83/526, Near D. Y. Patil Engg. College, Sant Tukaram Nagar, Pimpri, Pune - 411018.

© 09822136161 / 09975706701 / 07722092715 autocalsystemsmed@gmail.com / autocalsystems@gmail.com

www.autocalsystems.com



AUTOCAL SYSTEMS

Calibration of Biomedical, QA of Radiology, Thermal, Mechanical, Electrotechnical, Mass, Volume, Weights, Mapping and Validation Testing Services of Laminar Air Flow, Biosafety Cabinet, OT Validation, AHU and Calibration Facility in Lab / Onsite.



NABL ACCREDITED CALIBRATION LAB WITH ISO/IEC 17025

CALIBRATION CERTIFICATE

1. CUSTOMER
BANSAL HOSPITAL
SHAHPURA,
BHOPAL-462016.

Ambient Temp. :- (25 ± 4)°C
RH :- < 70%
Location of calibration :- ON SITE
Condition of Item :- OK

Page No. :- 1 of 1
Date of Issue :- 7/9/2020
Service Request No. :-
Certificate No. :- REF-02
Date of Calibration :- 7/9/2020
Next Calibration Due On :- 6/9/2021
Calibration method No. :- ACS/CM-101
Discipline :- Thermal Calibration
Parameter :- Temp

2. Description of Item

Name :- TEMPERATURE INDICATOR WITH SENSOR
Sr.No :- RT26FAJSASL/LT/2013
Make :- SAMSUNG
Model :- N.A
Type :- DIGITAL
Asset no :- N.A
Range :- 2 TO 8 °C
Resolution :- 1 °C
Specified Accuracy :- N.A
Location :- REFRIGERATOR
Department :- CENTRAL PHARMACY

3. Detail of Equipment used for calibration

Name :- Universal Calibrator
Make/I.D No. :- Massibus/ACS-EQP-06
Certificate No. :- CC/ECL/1442/19-20
Certified By :- IDEMI
Calibration Validity :- 13-Nov-20
PT-100 Sensor
Tempsens/ACS-EQP-05
CC/TRL/0607/19-20
IDEMI
17-Nov-20

4. Calibration Results

Calibration Points °C	Converted Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	1.7	2	0.3	0.50
4	3.7	4	0.3	0.50
6	5.7	6	0.3	0.50
7	6.7	7	0.3	0.50
8	7.7	8	0.3	0.50

Note:

- The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to coverage probability of approximately 95% for normal distribution.
- This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- Calibration point were selected as per customer specifications and Calibration report traceable to National or International Standard.
- This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Autocal Systems, Pune"

Calibrated By: *Aravali*
Calibration Engineer



Approved By: *Mr. Ankush Bagde*
Technical Manager
(Mr. Ankush Bagde)

83/526, Near D. Y. Patil Engg. College, Sant Tukaram Nagar, Pimpri, Pune - 411018.

09822136161 / 09975706701 / 07722092715 autocalsystemsmed@gmail.com / autocalsystems@gmail.com

www.autocalsystems.com



AUTOCAL SYSTEMS

Calibration of Biomedical, QA of Radiology, Thermal, Mechanical, Electrotechnical, Mass, Volume, Weights, Mapping and Validation Testing Services of Laminar Air Flow, Biosafety Cabinet, OT Validation, AHU and Calibration Facility in Lab / Onsite.



NABL ACCREDITED CALIBRATION LAB WITH ISO/IEC 17025

CALIBRATION CERTIFICATE

1. CUSTOMER :-
BANSAL HOSPITAL
SHAHPURA,
BHOPAL-462016.
Ambient Temp. :- (25 ± 4)°C
RH :- < 70%
Location of calibration :- ON SITE
Condition of Item :- OK

Page No. :- 1 of 1
Date of Issue :- 7/9/2020
Service Request No. :- --
Certificate No. :- REF-04
Date of Calibration :- 7/9/2020
Next Calibration Due On :- 8/9/2021
Calibration method No. :- ACS/CM-101
Discipline :- Thermal Calibration
Parameter :- Temp

2. Description of Item

Name :- TEMPERATURE INDICATOR WITH SENSOR Range :- 2 TO 8 °C
Sr.No :- 561007616B000259 Resolution :- 1 °C
Make :- SAMSUNG Specified Accuracy :- N.A
Model :- N.A Location :- REFRIGERATOR
Type :- DIGITAL Department :- CENTRAL PHARMACY
Asset no :- N.A

3. Detail of Equipment used for calibration


Name :- Universal Calibrator PT-100 Sensor
Make/I.D No. :- Massibus/ACS-EQP-06 Tempsens/ACS-EQP-05
Certificate No. :- CC/ECL/1442/19-20 CC/TRL/0607/19-20
Certified By :- IDEMI IDEMI
Calibration Validity :- 13-Nov-20 17-Nov-20


4. Calibration Results :-


Calibration Points °C	Converted Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	1.5	2	0.5	0.50
4	3.5	4	0.5	0.50
6	5.5	6	0.5	0.50
7	6.5	7	0.5	0.50
8	7.5	8	0.5	0.50

Note:

- The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to coverage probability of approximately 95% for normal distribution.
- This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- Calibration point were selected as per customer specifications and Calibration report traceable to National or International Standard.
- This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Autocal Systems, Pune"

Calibrated By: 
Calibration Engineer



Approved By: 
Technical Manager
(Mr. Ankush Bagde)



AUTOCAL SYSTEMS

Calibration of Biomedical, QA of Radiology, Thermal, Mechanical, Electrotechnical, Mass, Volume, Weights, Mapping and Validation Testing Services of Laminar Air Flow, Biosafety Cabinet, OT Validation, AHU and Calibration Facility in Lab / Onsite.



NABL ACCREDITED CALIBRATION LAB WITH ISO/IEC 17025

CALIBRATION CERTIFICATE

1. CUSTOMER
BANSAL HOSPITAL
SHAHPURA,
BHOPAL-462016.
Ambient Temp. :- (25 ± 4)°C
RH :- < 70%
Location of calibration :- ON SITE
Condition of Item :- OK

Page No. :- 1 of 1
Date of Issue :- 7/9/2020
Service Request No. :-
Certificate No. :- REF-03
Date of Calibration :- 7/9/2020
Next Calibration Due On :- 6/9/2021
Calibration method No. :- ACS/CM-101
Discipline :- Thermal Calibration
Parameter :- Temp

2. Description of Item

Name :- TEMPERATURE INDICATOR WITH SENSOR
Sr.No :- 53024DBD800191Y
Make :- VOLTAS
Model :- N.A
Type :- DIGITAL
Asset no :- N.A
Range :- 2 TO 8 °C
Resolution :- 1 °C
Specified Accuracy :- N.A
Location :- REFRIGERATOR
Department :- CENTRAL PHARMACY

3. Detail of Equipment used for calibration

Name :- Universal Calibrator
Make/I.D No. :- Massibus/ACS-EQP-06
Certificate No. :- CC/ECL/1442/19-20
Certified By :- IDEMI
Calibration Validity :- 13-Nov-20
PT-100 Sensor
Tempsens/ACS-EQP-05
CC/TRL/0607/19-20
IDEMI
17-Nov-20


4. Calibration Results

Calibration Points °C	Converted Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	1.3	2	0.7	0.50
4	3.3	4	0.7	0.50
6	5.3	6	0.7	0.50
7	6.3	7	0.7	0.50
8	7.3	8	0.7	0.50

Note:

- 1) The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to coverage probability of approximately 95% for normal distribution.
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications and Calibration report traceable to National or International Standard.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Autocal Systems, Pune"

Calibrated By: *Arushi*
Calibration Engineer



Approved By: *Arushi Bagde*
Technical Manager
(Mr. Ankush Bagde)

83/526, Near D. Y. Patil Engg. College, Sant Tukaram Nagar, Pimpri, Pune - 411018.

☎ 09822136161 / 09975706701 / 07722092715 ✉ autocalsystemsmed@gmail.com / autocalsystems@gmail.com

www.autocalsystems.com



AUTOCAL SYSTEMS
 Calibration of Biomedical, QA of Radiology, Thermal, Mechanical, Electrotechnical,
 Mass, Volume, Weights, Mapping and Validation Testing Services of Laminar Air Flow,
 Biosafety Cabinet, OT Validation, AHU and Calibration Facility in Lab / Onsite.



NABL ACCREDITED CALIBRATION LAB WITH ISO/IEC 17025

CALIBRATION CERTIFICATE

1. CUSTOMER :-
BANSAL HOSPITAL
 SHAHPURA,
 BHOPAL-462016.
 Ambient Temp. :- (25 ± 4)°C
 RH :- < 70%
 Location of calibration :- ON SITE
 Condition of Item :- OK

Page No. :- 1 of 1
 Date of Issue :- 7/9/2020
 Service Request No. :-
 Certificate No. :- REF-15
 Date of Calibration :- 7/9/2020
 Next Calibration Due On :- 6/9/2021
 Calibration method No. :- ACS/CM-101
 Discipline :- Thermal Calibration
 Parameter :- Temp

2. Description of Item

Name :- TEMPERATURE INDICATOR WITH SENSOR
 Sr.No :- 131253912
 Make :- CELFROST
 Model :- N.A
 Type :- DIGITAL
 Asset no :- N.A
 Range :- 2 TO 8 °C
 Resolution :- 1 °C
 Specified Accuracy :- N.A
 Location :- REFRIGERATOR
 Department :- IPD PHARMACY

3. Detail of Equipment used for calibration

Name :- Universal Calibrator
 Make/ID No. :- Massibus/ACS-EQP-06
 Certificate No. :- CC/ECU/1442/19-20
 Certified By :- IDEMI
 Calibration Validity :- 13-Nov-20
 PT-100 Sensor
 Tempsens/ACS-EQP-05
 CC/TRU/0607/19-20
 IDEMI
 17-Nov-20


4. Calibration Results :-

Calibration Points °C	Converted Standard Reading °C	UUC Reading °C	Error in °C	Expanded Uncertainty in ± °C
2	1.3	2	0.7	0.50
4	3.3	4	0.7	0.50
6	5.3	6	0.7	0.50
7	6.3	7	0.7	0.50
8	7.3	8	0.7	0.50

Note:

- The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor k=2, which corresponds to coverage probability of approximately 95% for normal distribution.
- This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- Calibration point were selected as per customer specifications and Calibration report traceable to National or International Standard.
- This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Autocal Systems, Pune"

Calibrated By *Anjali*
 Calibration Engineer



Approved By *Ankush Bagde*
 Technical Manager
 (Mr. Ankush Bagde)