



Temperature Check	Reaction Tray 37 Celsius		
Cuvette Check	Gain	ADC	Range
Wave Length			
340 nm			
380 nm			
404 nm			
416 nm			
450 nm			
476 nm			
500 nm			
524 nm			
548 nm			
572 nm			
604 nm			
628 nm			
660 nm			
700 nm			
748 nm			
804 nm			

This is to certify that the above mentioned product has been calibrated according to the standard operating procedure given by the Toshiba Medical Systems Corporation, Japan(Doc #2B586-297EN)

Calibrator used Multicalibrator Lot No.: 38120067 Exp.: 1/2020
Control used Norm Lot No.: 26471 Exp.: 28/2/2023
 Path Lot No.: 26472 Exp.: 28/2/2023
Test CV%
 TP 0.8% Date of Calibration: 15/Sep/2020
 GOT 0.6% Calibration Validity: 14/March/2021

For AGAPPE DIAGNOSTICS LTD.



Authorized Signatory
 Name: Richard Basil
 Designation: Engineer-Customer Service

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
 Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India
 Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
 X-508-CB, Block No. 32, KINFRA Small Industrial Park,
 Nerul, Cochin, Kerala, India - 685 721. Tel: +91 484 230 7477.

KOLKATA OFFICE
 406, Newlin Matrix, Plot No-10, Block-D4, Sector V,
 Salt Lake City, Kolkata - 700 091.
 Tel: +91 33 4003 0450 | Email: kolkataoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
 403 & 402, 4th Floor, Shivajinagar Business Centre, D9,
 Sakar Road, Parsipada, Andheri (East), Mumbai - 400 099, India.
 Tel: +91 22 4300 8000 | Email: mumbai@agappe.in

DELHI OFFICE
 D5M 540, 5th Floor, DLF Tower, Shrawang Marg,
 New Delhi - 110 014, India.
 Tel: +91 11 4556 8416 | Email: delhi@agappe.in

BANGALORE OFFICE
 5-5, B¹ Floor, Red Cross Bhawan, No.26 Race Course Road,
 Bangalore - 560 001. Tel: +91 80 2228 8288
 Email: bangaloreoffice@agappe.in

**CERTIFICATE OF CALIBRATION**

ADL/BO/SER/085/20-21

DATE : 19-01-2021	SITE : Sparsha Diagnostics, Mangalore	CITY : Mangalore
STATE : KARNATAKA	COUNTRY : INDIA	S/N. : AAA15X2088
INSTRUMENT NAME	TBA25FR	

Section	Work	Clean / Check	Lubricate	Replace
Reaction Tray	Cleaning of Reaction try	<input type="checkbox"/> <input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Sampling Pump	Cleaning	<input type="checkbox"/> <input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Reagent Pump	Cleaning	<input type="checkbox"/> <input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Probe Washing Pump	Cleaning	<input type="checkbox"/> <input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Reaction Cuvette washing	Cleaning	<input type="checkbox"/> <input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Sample Probe	Cleaning	<input type="checkbox"/> <input type="checkbox"/>		<input type="checkbox"/>
Reagent Probe	Cleaning	<input type="checkbox"/> <input type="checkbox"/>		<input type="checkbox"/>
Movement Check				
Reaction Tray	Checking of Cooling Temp.	<input type="checkbox"/>		<input type="checkbox"/>
Sampling Pump	Movement Check	<input type="checkbox"/>		<input type="checkbox"/>
Reagent pump	Movement Check	<input type="checkbox"/>		<input type="checkbox"/>
Probe washing Check	Movement Check	<input type="checkbox"/>		<input type="checkbox"/>
Reaction cuvette washing	Movement Check	<input type="checkbox"/>		<input type="checkbox"/>
Sampling probe	Movement Check	<input type="checkbox"/>		<input type="checkbox"/>
Reagent probe	Movement Check	<input type="checkbox"/>		<input type="checkbox"/>
Rack Detector	Function Check	<input type="checkbox"/>		<input type="checkbox"/>
Automated Sampler	Checking on reading of Sample Rack	<input type="checkbox"/>		<input type="checkbox"/>
Stirrer Check	Up & Down Movement, Rotation	<input type="checkbox"/>		<input type="checkbox"/>
Water supply & Drainage	Movement Check on Drainage pump	<input type="checkbox"/>		<input type="checkbox"/>
Body	Checking of leakage	<input type="checkbox"/>		<input type="checkbox"/>
Total Inspection	Cell blank Test	<input type="checkbox"/>		<input type="checkbox"/>

Authorized Signatory:**AGAPPE DIAGNOSTICS LTD.**

ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Postnattam (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
K/568-CB, Block No. 32, KMFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
404, Berlin Metro, Plot No-10, Block-D1, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI [REGISTERED OFFICE]
401 & 402, 4th Floor, Sansingh Business Centre, D9,
Satya Road, Worliwadi, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
D5M 540, 5th Floor, DLF Tower, Shrawj Marg,
New Delhi - 110 015, India.
Tel: +91 11 4358 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, 1st Floor, Red Cross Bhawan, No.25 Race Course Road,
Bangalore - 560 001, Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

YOUR BEST PARTNER IN DIAGNOSTICS