

TRANSASIA®

Transasia Bio-Medicals Ltd., No. 43, Castle Street, Richmond Road, Opp. Empire Suitess, Ashalnagar Bengaluru - 560025
Tel: 080 4430 9000 / 2556 8044 Email : sz2@transasia.co.in CIN : U33110MH1985PLC036198



Date: 25-06-2020
Effective Date: 25-06-2020

Certificate of Calibration

Customer Name: SPARSHA DIAGNOSTICS , MANGALORE.

Model : Automated Hematology Analyzer Sysmex XNL-330

Serial No. : 11093

Calibration Done Date: 25.6.20

Next Calibration Due Date On or Before: 24-07-2020

Lab In-charge: . RANI D SOZA

This is to certify that the above-mentioned product has been verified of calibration according to the standard procedures provided by Sysmex Corporation, Japan.

The reference instruments used for value-assignment are managed by the traceability system in Sysmex Corporation and these are traceable to the International Standards, such as ICSH.



Calibration at site performed by
Ravi Chavan
DY RSM
Transasia Bio-Medicals Ltd
Mumbai

Encl:

1. Certificate of Inspection
2. Assay Sheet of Calibrator XN
3. Printouts
4. Traceability & Uncertainty document

www.transasia.co.in



HO : Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai - 400 072 Tel: +91 22 4030 9000 Fax: +91 22 2857 3030

Transasia Bio-Medicals Ltd., No. 43, Castle Street, Richmond Road, Opp. Empire Suitess, Ashalnagar Bengaluru - 560025
Tel: 080 4430 9000 / 2556 8044 Email : sz2@transasia.co.in CIN : U33110MH1985PLC036198

TRANSASIA[®]



UNMATCHED SERVICE
25-06-~~SINCE 1979~~...
25-06-2020

Date:
Effective Date:

Certificate of Inspection

1. Model: Automated Hematology Analyzer Sysmex XN1-330
2. Serial No.: 11093
3. Calibration Date: 25-06-2020
4. Material used: XN-CAL (Lot No. 0167 2101, Expiry date: 19-July-2020)

By comparing your data to the results of the standard counters in Sysmex Corporation, the calibration for XN Analyzer parameters using the measurement standard material (XN-CAL) was completed. The calibration result of 10 runs is summarized in the following table, Please refer to the attached sheets for the details.



Technical Service Department
Transasia Bio-Medicals Ltd

www.transasia.co.in



HO : Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai - 400 072 Tel: +91 22 4030 9000 Fax: +91 22 2857 3030



5. BACKGROUND CHECK

PARAMETER	RESULT	Range
WBC-N	0	$0.1 \cdot 10^3$ uL or less
WBC-D	0	$0.1 \cdot 10^3$ uL or less
RBC	0	$0.02 \cdot 10^{12}$ uL or less
HGB	0	0.1g/dl or less
PLT	0	$10 \cdot 10^6$ uL or less



Technical Service Department
Transasia Bio-Medicals Ltd