



# ADARSHA CALIBRATION SERVICES

**NABL ACCREDITED CALIBRATION LABORATORY**

Plot No.18,19, Survey No-118, Byraveshwara Nagar, Manganahalli Cross, Ullal Main Road,  
Ullal (P), Bangalore-560056, Ph: +91-80-2324 0578, Mob: +91-9945086171/172/173

E: [acscalib@gmail.com](mailto:acscalib@gmail.com) / [enquiry@adarshacalibrations.com](mailto:enquiry@adarshacalibrations.com) W: [www.adarshacalibrations.com](http://www.adarshacalibrations.com)



CC-2722

## CERTIFICATE OF CALIBRATION

Page 1 of 1

Certificate Number: ACS/NOV/20/S15-05		ULR Number : ULR-CC-2722000005573F	
Customer Name & Address: M/s. Sparsha Diagnostics, Mangalore.			
Date Of SRF: 16-11-2020	Calibrated At: Site		
Date of Calibration: 16-11-2020	Calibration WI No: WI/TH/04		
Next Cal Due: 15-05-2021	Date of Issue: 23-11-2020		
Environmental Conditions: Temperature: 25.8°C		Humidity: 51% R.H	
DETAILS OF DEVICE UNDER CALIBRATION			
Nomenclature : Refrigerator	Serial. No: QAECG1100A170601205		
Make & Model : Elam Pro & NA	Code/ID No : SD/CL/REF/04		
Range : 2 to 8°C	Resolution: NA		
Location: NA			
DETAILS OF STANDARD INSTRUMENT USED		Reference Standard: DKD-R-5-7	
1.Nomenclature: Digital Thermometer with Sensor	Serial.No/Id.No: 13021909/ACS-T/05		
Make / Model : YCT & RTD-116	Cal Due Date: 20-05-2021		
Range : -80 to 400 °C	Certificate No: ACS/TH/20-21/102425-01		
CALIBRATION RESULTS:			
Parameter: Temperature		Discipline : Thermal	
Sl No	Operating Range in °C	Measured Value in °C	Expanded Measurement Uncertainty ± in °C
Refrigerator			
1	2 to 8	5.12	0.50
REMARKS & CONCLUSION:			
<p>a. The Measurement Uncertainty is estimated at a confidence level of 95.45 % with a coverage factor k = 2.00.</p> <p>b. Next calibration due date mentioned as per customer requirement</p> <p>c. Calibration work order received through M/s.Medical Engineering &amp; Services.</p>			
Calibrated By		Authorized Signatory	
Deepak Kamaraj Calibration Engineer		Manjunath I Technical Manager	

