



ADARSHA CALIBRATION SERVICES

NABL ACCREDITED CALIBRATION LABORATORY

Plot No.18,19, Survey No-118, Byraveshwara Nagar, Manganahalli Cross, Ullal Main Road,
Ullal (P), Bangalore-560056, Ph: +91-80-2324 0578, Mob: +91-9945086171/172/173

E: acslib@gmail.com / enquiry@adarshacalibrations.com W: www.adarshacalibrations.com



CC-2722

CERTIFICATE OF CALIBRATION

Page 1 of 1

Certificate Number:	ACS/NOV/20/S15-06	ULR Number :	ULR-CC-2722000005574F
Customer Name & Address:	M/s. Sparsha Diagnostics, Mangalore.		
Date Of SRF:	16-11-2020	Calibrated At:	Site
Date of Calibration:	16-11-2020	Calibration WI No:	WI/TH/04
Next Cal Due:	15-05-2021	Date of Issue:	23-11-2020
Environmental Conditions:	Temperature: 25.4°C	Humidity:	52% R.H
DETAILS OF DEVICE UNDER CALIBRATION			
Nomenclature :	Refrigerator	Serial. No:	A29053719
Make & Model :	Cel Frost & NA	Code/ID No :	SD/CL/REF/05
Range :	2 to 8°C	Resolution:	NA
Location:	NA		
DETAILS OF STANDARD INSTRUMENT USED			
1.Nomenclature:	Digital Thermometer with Sensor	Reference Standard:	DKD-R-5-7
Make / Model :	YCT & RTD-116	Serial.No/Id.No:	13021909/ACS-T/05
Range :	-80 to 400 °C	Cal Due Date:	20-05-2021
		Certificate No:	ACS/TH/20-21/102425-01
CALIBRATION RESULTS:			
Parameter: Temperature			Discipline : Thermal
Sl No	Operating Range in °C	Measured Value in °C	Expanded Measurement Uncertainty ± in °C
Refrigerator			
1	2 to 8	5.74	0.50
REMARKS & CONCLUSION:			
<p>a. The Measurement Uncertainty is estimated at a confidence level of 95.45 % with a coverage factor $k = 2.00$.</p> <p>b. Next calibration due date mentioned as per customer requirement</p> <p>c. Calibration work order received through M/s.Medical Engineering & Services.</p>			
Calibrated By		Authorized Signatory	
 Deepak Kamaraj Calibration Engineer		 Manjunath I Technical Manager	
			