

1

CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/20-21/00433

Page: 1 of 2

Customer Name:	DIA CARE DIAGNOSTICS	Customer Address:	#2/751, VEPPANAPALLI, KRISHNAGIRI DT, TAMIL NADU
Location:	BIO CHEMISTRY	Serial No:	40289
Description:	BIO CHEMISTRY ANALYSER	Tag/Asset No:	NA
Make:	TULIP	Date of Calibration:	25.01.2021
Model:	EVOLUTION 3000	Date of Due:	24.01.2021

Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1	Electrical Safety Analyzer	Fluke	ESA615	2244202	26-Jun-2021

Environmental Conditions and Standard Operating Procedure Details:

Temperature	24°C	Calibration Done At	ONSITE
6 · · · ·			CP-BM001-2019/ES
Humidity (40-70% RH)	59%	Calibration Procedure No	

REMARKS:

The above mentioned instruments has been calibrated using standard manufacturer recommended protocols, using equipments having traceability to National/International Standards. Partial reproduction of the certificates is not permitted.

Calibrated by:

Authorized Signatory:

(Calibration Engineer)

(Technical Manager)



Page: 2 of 2

Electrical Safety Test Passed : Yes / No

	Electrical Safety Test	T
Parameters Used	Measured in Test Gadget	Limits
Voltage N-E	2.3 V	< 3V
Voltage P-N	220 V	< 230V
Voltage P-E	222 V	< 230V
Maximum Current Taken by the DUT	0.1	As per Manufacture specification
Earth Leakage Current NC	121 μΑ	<500μA for B,BF,CF
Earth Leakage Current SFC	237 μΑ	<1000µA for B,BF,CF
Enclosure Leakage Current NC	5.1μΑ	<100μA for B,BF,CF
Enclosure Leakage Current SFC	5.3 μΑ	<500μA for B,BF,CF
Insulation Resistance M-P.E	Good	More than 100Mohm
Insulation Resistance M-A.P	Good	More than 100Mohm
Insulation Resistance A.P-P.E	Good	More than 100Mohm

CALIBRATION RESULT:	
	Corrective maintenance required
s	Removed from use
<u> </u>	Acceptable for use
Calibration Engineer Sign	00

Date: 25-01-2021