

LAB MONTHLY SUMMARY



Lab Name **MOLXPRT DIAGNOSTICS PVT LTD** Lab No **14224**
 Month **January** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 20/01/2021

Date of Report Published : 05/02/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Biosystems	255	119.68	9.01	10.79	117 mg/dl	-0.25	1.35
2	UREA II	Urease UV / GLDH II	Biosystems	193	31.36	10.37	3.25	31 mg/dl	-0.11	0.47
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Biosystems	188	1.37	11.10	0.15	1.4 mg/dl	0.20	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Biosystems	213	2.27	21.23	0.48	2.2 mg/dl	-0.15	0.07
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	209	5.46	10.49	0.57	4.9 g/dl	-0.98	0.08
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	204	3.33	9.62	0.32	3 g/dl	-1.03	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	202	5.61	14.10	0.79	5.1 mg/dl	-0.64	0.11
8	CHOLESTROL II	CHOD-PAP II	Biosystems	240	115.43	10.70	12.35	106 mg/dl	-0.76	1.59
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	208	143.79	12.33	17.73	139 mg/dl	-0.27	2.46

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****