

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE

|                          |                                                                                                      |                 |                  |
|--------------------------|------------------------------------------------------------------------------------------------------|-----------------|------------------|
| Customer Name & Address: | M/S. Sobhi Diagnostic Centre,<br>18/4, Perumal Coil street,<br>Vallioor – 627 117, Tirunelveli Dist. | Certificate No: | AMS/HOSP/2889150 |
|                          |                                                                                                      | Issuing Date:   | 27/01/2021       |
|                          |                                                                                                      | Page No:        | 1 of 1           |

### Details of Device under Calibration (DUC):

|                      |                  |                            |                  |
|----------------------|------------------|----------------------------|------------------|
| DUC:                 | Micro Pipette    | Id. No:                    | SDCV/LAB/VMP/001 |
| Make:                | Lab Auto Pipette | Department:                | Laboratory       |
| Model:               | Variable         | Calibrated On:             | 26/01/2021       |
| Sl. No:              | ---              | Calibration Due Suggested: | 25/01/2022       |
| Type:                | Variable         | DUC Condition on Receipt:  | OK               |
| Range & Least Count: | 5 to 50µl        |                            |                  |

### Traceability Details:

| Sl.No. | Test Equipment            | Make       | Certificate Number | Valid Up to |
|--------|---------------------------|------------|--------------------|-------------|
| 1      | Electronic Weighing Scale | EWC/SF400C | LM-2008863         | 08/08/2021  |
| 2      | Thermo Hygrometer         | HTC        | QC/AMS-DTI-004     | 22/02/2021  |

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

### Test Results:

| SL. No. | DUC Reading (µl) | Standard Reading (µl) |         | Correction (µl) |
|---------|------------------|-----------------------|---------|-----------------|
|         |                  | Trial 1               | Trial 2 |                 |
| 1       | 5                | 5                     | 5       | 0               |
| 2       | 10               | 10                    | 10      | 0               |
| 3       | 20               | 20                    | 20      | 0               |
| 4       | 50               | 50                    | 50      | 0               |

Allowable Tolerance: ±1%

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)

B. S. S.



### Approved by:

(Technical Manager)

B. S. S.

aceagm@gmail.com

www.acemedical.in

93455 56140, 93455 56148

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE

|                          |                                                                                                      |                 |                  |
|--------------------------|------------------------------------------------------------------------------------------------------|-----------------|------------------|
| Customer Name & Address: | M/S. Sobhi Diagnostic Centre,<br>18/4, Perumal Coil street,<br>Vallioor – 627 117, Tirunelveli Dist. | Certificate No: | AMS/HOSP/2889151 |
|                          |                                                                                                      | Issuing Date:   | 27/01/2021       |
|                          |                                                                                                      | Page No:        | 1 of 1           |

### Details of Device under Calibration (DUC):

|                      |               |                            |                  |
|----------------------|---------------|----------------------------|------------------|
| DUC:                 | Micro Pipette | Id. No:                    | SDCV/LAB/VMP/002 |
| Make:                | Lab Serv      | Department:                | Laboratory       |
| Model:               | Variable      | Calibrated On:             | 26/01/2021       |
| Sl. No:              | ZY 63888      | Calibration Due Suggested: | 25/01/2022       |
| Type:                | Variable      | DUC Condition on Receipt:  | OK               |
| Range & Least Count: | 5 to 50µl     |                            |                  |

### Traceability Details:

| Sl.No. | Test Equipment            | Make       | Certificate Number | Valid Up to |
|--------|---------------------------|------------|--------------------|-------------|
| 1      | Electronic Weighing Scale | EWC/SF400C | LM-2008863         | 08/08/2021  |
| 2      | Thermo Hygrometer         | HTC        | QC/AMS-DTI-004     | 22/02/2021  |

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

### Test Results:

| SL. No. | DUC Reading (µl) | Standard Reading (µl) |         | Correction (µl) |
|---------|------------------|-----------------------|---------|-----------------|
|         |                  | Trial 1               | Trial 2 |                 |
| 1       | 5                | 5                     | 5       | 0               |
| 2       | 10               | 10                    | 10      | 0               |
| 3       | 20               | 20                    | 20      | 0               |
| 4       | 50               | 50                    | 50      | 0               |

Allowable Tolerance: ±1%

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)

B. S. S.



aceagm@gmail.com

www.acemedical.in



### Approved by:

(Technical Manager)

B. S. S.

93455 56140, 93455 56148

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE

|                          |                                                                                                      |                 |                  |
|--------------------------|------------------------------------------------------------------------------------------------------|-----------------|------------------|
| Customer Name & Address: | M/S. Sobhi Diagnostic Centre,<br>18/4, Perumal Coil street,<br>Vallioor – 627 117, Tirunelveli Dist. | Certificate No: | AMS/HOSP/2889148 |
|                          |                                                                                                      | Issuing Date:   | 27/01/2021       |
|                          |                                                                                                      | Page No:        | 1 of 1           |

### Details of Device under Calibration (DUC):

|                      |                  |                            |                  |
|----------------------|------------------|----------------------------|------------------|
| DUC:                 | Micro Pipette    | Id. No:                    | SDCV/LAB/FMP/001 |
| Make:                | Lab Micropipette | Department:                | Laboratory       |
| Model:               | Fixed            | Calibrated On:             | 26/01/2021       |
| Sl. No:              | ---              | Calibration Due Suggested: | 25/01/2022       |
| Type:                | Variable         | DUC Condition on Receipt:  | OK               |
| Range & Least Count: | 100µl            |                            |                  |

### Traceability Details:

| Sl.No. | Test Equipment            | Make       | Certificate Number | Valid Up to |
|--------|---------------------------|------------|--------------------|-------------|
| 1      | Electronic Weighing Scale | EWC/SF400C | LM-2008863         | 08/08/2021  |
| 2      | Thermo Hygrometer         | HTC        | QC/AMS-DTI-004     | 22/02/2021  |

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

### Test Results:

| SL. No. | DUC Reading (µl) | Standard Reading (µl) |         | Correction (µl) |
|---------|------------------|-----------------------|---------|-----------------|
|         |                  | Trial 1               | Trial 2 |                 |
| 1       | 100              | 100                   | 100     | 0               |

Allowable Tolerance: ±1%

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)

B. S. S.



### Approved by:

(Technical Manager)

B. S. S.

aceagm@gmail.com

www.acemedical.in

93455 56140, 93455 56148

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE

|                          |                                                                                                      |                 |                  |
|--------------------------|------------------------------------------------------------------------------------------------------|-----------------|------------------|
| Customer Name & Address: | M/S. Sobhi Diagnostic Centre,<br>18/4, Perumal Coil street,<br>Vallioor – 627 117, Tirunelveli Dist. | Certificate No: | AMS/HOSP/2889152 |
|                          |                                                                                                      | Issuing Date:   | 27/01/2021       |
|                          |                                                                                                      | Page No:        | 1 of 1           |

### Details of Device under Calibration (DUC):

|                      |               |                            |                  |
|----------------------|---------------|----------------------------|------------------|
| DUC:                 | Micro Pipette | Id. No:                    | SDCV/LAB/VMP/003 |
| Make:                | ASTRA         | Department:                | Laboratory       |
| Model:               | AMP 103LX     | Calibrated On:             | 26/01/2021       |
| Sl. No:              | YE3K 103225   | Calibration Due Suggested: | 25/01/2022       |
| Type:                | Variable      | DUC Condition on Receipt:  | OK               |
| Range & Least Count: | 100 to 1000µl |                            |                  |

### Traceability Details:

| Sl.No. | Test Equipment            | Make       | Certificate Number | Valid Up to |
|--------|---------------------------|------------|--------------------|-------------|
| 1      | Electronic Weighing Scale | EWC/SF400C | LM-2008863         | 08/08/2021  |
| 2      | Thermo Hygrometer         | HTC        | QC/AMS-DTI-004     | 22/02/2021  |

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

### Test Results:

| SL. No. | DUC Reading (µl) | Standard Reading (µl) |         | Correction (µl) |
|---------|------------------|-----------------------|---------|-----------------|
|         |                  | Trial 1               | Trial 2 |                 |
| 1       | 100              | 100                   | 100     | 0               |
| 2       | 200              | 200                   | 200     | 0               |
| 3       | 500              | 500                   | 500     | 0               |
| 4       | 1000             | 1000                  | 1000    | 0               |

Allowable Tolerance: ±1%

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)

B. S. S.



### Approved by:

(Technical Manager)

B. S. S.

aceagm@gmail.com

www.acemedical.in

93455 56140, 93455 56148

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE



|                          |                                                                                                      |                 |                  |
|--------------------------|------------------------------------------------------------------------------------------------------|-----------------|------------------|
| Customer Name & Address: | M/S. Sobhi Diagnostic Centre,<br>18/4, Perumal Coil street,<br>Vallioor – 627 117, Tirunelveli Dist. | Certificate No: | AMS/HOSP/2889149 |
|                          |                                                                                                      | Issuing Date:   | 27/01/2021       |
|                          |                                                                                                      | Page No:        | 1 of 1           |

### Details of Device under Calibration (DUC):

|                      |                  |                            |                  |
|----------------------|------------------|----------------------------|------------------|
| DUC:                 | Micro Pipette    | Id. No:                    | SDCV/LAB/FMP/002 |
| Make:                | Lab Micropipette | Department:                | Laboratory       |
| Model:               | Fixed            | Calibrated On:             | 26/01/2021       |
| Sl. No:              | ---              | Calibration Due Suggested: | 25/01/2022       |
| Type:                | Variable         | DUC Condition on Receipt:  | OK               |
| Range & Least Count: | 500µl            |                            |                  |

### Traceability Details:

| Sl.No. | Test Equipment            | Make       | Certificate Number | Valid Up to |
|--------|---------------------------|------------|--------------------|-------------|
| 1      | Electronic Weighing Scale | EWC/SF400C | LM-2008863         | 08/08/2021  |
| 2      | Thermo Hygrometer         | HTC        | QC/AMS-DTI-004     | 22/02/2021  |

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

### Test Results:

| SL. No. | DUC Reading (µl) | Standard Reading (µl) |         | Correction (µl) |
|---------|------------------|-----------------------|---------|-----------------|
|         |                  | Trial 1               | Trial 2 |                 |
| 1       | 500              | 500                   | 500     | 0               |

Allowable Tolerance: ±1%

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)

*B. S. S.*



### Approved by:

(Technical Manager)

*B. S. S.*