

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Sobhi Diagnostic Centre, 18/4, Perumal Coil street, Vallioor – 627 117, Tirunelveli Dist.	Certificate No:	AMS/LAB/2889154
		Issuing Date:	27/01/2021
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Refrigerator – Reagent	Id. No:	SDCV/LAB/REFR/001
Make:	LG	Department:	Laboratory
Model:	GL B281 BPZX/2019	Calibrated On:	26/01/2021
Sl. No:	901PRAW003688	Calibration Due Suggested:	25/01/2022
Type:	Reagents	DUC Condition on Receipt:	OK
Range & Least Count:	0 – 8°C / LC 1°C		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	QC/AMS/DTI/002	22/02/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

### Test Results:

SL.No.	Place of Reading	Standard Reading (°C)	Correction (°C)
1	Chiller	-0.87	0
2	Chamber	4.23	0

Allowable Tolerance: User Defined

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer).

B. S. S.



### Approved by:

(Technical Manager)

B. S. S.

aceagm@gmail.com

www.acemedical.in

93455 56140, 93455 56148

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Sobhi Diagnostic Centre, 18/4, Perumal Coil street, Vallioor – 627 117, Tirunelveli Dist.	Certificate No:	AMS/LAB/2889155
		Issuing Date:	27/01/2021
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Refrigerator – Samples	Id. No:	SDCV/LAB/REFS/001
Make:	Electrolux	Department:	Laboratory
Model:	Small	Calibrated On:	26/01/2021
Sl. No:	---	Calibration Due Suggested:	25/01/2022
Type:	Blood Samples	DUC Condition on Receipt:	OK
Range & Least Count:	0 – 8°C / LC 1°C		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	QC/AMS/DTI/002	22/02/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

### Test Results:

SL.No.	Place of Reading	Standard Reading (°C)	Correction (°C)
1	Chiller	1.23	0
2	Chamber	7.09	0

Allowable Tolerance: User Defined

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:  
(Calibration Engineer)

B. S. S.



Approved by:  
(Technical Manager)

B. S. S.

aceagm@gmail.com

www.acemedical.in

93455 56140, 93455 56148