



CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/21-22/00963

Page: 1 of 2

Customer Name:	VIGNESH DIAGNOSTIC CENTRE	Customer Address:	79, REDHILLS HIGH ROAD, PUDUR, AMBATTUR, CHENNAI- 600053.
Location:	LAB	SRF No:	SBS/SRF/21-22/0017/001
Description:	MICRO PIPETTE	SRF Date:	11.02.2021
Make:	BIOHIT	Receipt Date:	11.02.2021
Model:	NA	Date of Calibration:	11.02.2021
Serial No:	YE6A729133	Date of Due:	10.02.2022
Range:	5-50µl (VARIABLE)	Issue Date:	11.02.2021

Traceability Details:

S.No	Standard Used	Make	Model	Cer.No	Validity
1	Electronic Weighing Balance	A&D	GH-252	ECT 2020/0914-01	14-Dec-2022

Environmental Conditions and Standard Operating Procedure Details:

Temperature	24°C	Calibration Done At	LAB
Humidity (40-70% RH)	55%	Calibration Procedure No	SBS-MBV-03

REMARKS:

Calibration Report : Attached

Calibrated by:

P. MYILSAMY

(Calibration Engineer)

Authorized Signatory:

C. SHANMUGARAJ
(Technical Manager)






Result of Calibration:

Page: 2 of 2

REPORT				
S.No	Units	Nominal Values	Measured Values	Tolerance
1	μL	10	10.0	0.5%
2	μL	20	20.0	0.5%
3	μL	30	30.1	0.5%
4	μL	40	40.2	0.5%
5	μL	50	50.1	0.5%

CALIBRATION RESULT:

- Corrective maintenance required
 Removed from use
 Acceptable for use

Calibration Engineer Sign: 

Date: 11.02.2021

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First floor , Poomagal Main Road, Ekkaltuthangal, Chennai - 600 032, Tel: 044 - 2225 2087.