



## CALIBRATION REPORT

### Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/21-22/00964

Page: 1 of 2

Customer Name:	VIGNESH DIAGNOSTIC CENTRE	Customer Address:	79, REDHILLS HIGH ROAD, PUDUR, AMBATTUR, CHENNAI- 600053.
Location:	LAB	SRF No:	SBS/SRF/21-22/0017/002
Description:	MICRO PIPETTE	SRF Date:	11.02.2021
Make:	BIOHIT	Receipt Date:	11.02.2021
Model:	NA	Date of Calibration:	11.02.2021
Serial No:	YE6F732383	Date of Due:	10.02.2022
Range:	100-1000µl (VARIABLE)	Issue Date:	11.02.2021

### Traceability Details:

S.No	Standard Used	Make	Model	Cer.No	Validity
1	Electronic Weighing Balance	A&D	GH-252	ECT 2020/0914-01	14-Dec-2022


### Environmental Conditions and Standard Operating Procedure Details:

Temperature	24°C	Calibration Done At	LAB
Humidity (40-70% RH)	55%	Calibration Procedure No	SBS-MBV-03

### REMARKS:

Calibration Report : Attached

Calibrated by:

  
P. MYILSAMY  
(Calibration Engineer)

Authorized Signatory:

  
C. SHANMUGARAJ  
(Technical Manager)



Result of Calibration:

Page: 2 of 2

REPORT				
S.No	Units	Nominal Values	Measured Values	Tolerance
1	μL	100	100.1	0.5%
2	μL	200	200.2	0.5%
3	μL	500	500.1	0.5%
4	μL	800	800.4	0.5%
5	μL	1000	1000.3	0.5%

CALIBRATION RESULT:

- Corrective maintenance required  
 Removed from use  
 Acceptable for use

Calibration Engineer Sign:

Date: 11.02.2021

**SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First floor , Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087.