

## CALIBRATION REPORT

Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/21-22/00965

Page: 1 of 2

Customer Name:	VIGNESH DIAGNOSTIC CENTRE	Customer Address:	79, REDHILLS HIGH ROAD, PUDUR, AMBATTUR, CHENNAI- 600053.
Location:	LAB	SRF No:	SBS/SRF/21-22/0017/003
Description:	ELECTRONIC WEIGHING BALANCE	SRF Date:	11.02.2021
Make:	INTEL	Receipt Date:	11.02.2021
Model:	DIGITAL WEIGHING SCALE	Date of Calibration:	11.02.2021
Serial No:	NA	Date of Due:	10.02.2022
Asset No:	NA	Issue Date:	11.02.2021

Traceability Details:

S.No	Standard Used	Make	Model	Serial No	Validity
1	Standard Weights	Sensortron	1Kg	033	12-Aug-2021
2	Standard Weights	Sensortron	2Kg	034	12-Aug-2021
3	Standard Weights	Sensortron	5 Kg	305	31-Mar-2021
1	Standard Weights	Sensortron	100g	2003068	10-Mar-2022
2	Standard Weights	Sensortron	200g	2003067	10- Mar -2022
3	Standard Weights	Sensortron	200g	2003066	10- Mar -2022


Environmental Conditions and Standard Operating Procedure Details:

Temperature	24°C	Calibration Done At	LAB
Humidity (40-70% RH)	55%	Calibration Procedure No	SBS-MBV-01


**REMARKS:**

The above mentioned instruments has been calibrated using standard manufacturer recommended protocols, using equipments having traceability to National/International Standards. Partial reproduction of the certificates is not permitted.

Calibrated by:

  
**P. MYILSAMY**  
 (Calibration Engineer)

Authorized Signatory:

  
**C. SHANMUGARAJU**  
 (Technical Manager)




**Result of Calibration:**

QUANTITATIVE TASKS					
S.No	Description	Units	Set Values	Measured Values	Limit/Tolerance
1.	Weight	g	100	100.1	±0.5
2.	Weight	g	200	200.1	±0.5
3.	Weight	g	500	500.2	±0.5
4.	Weight	Kg	1	1.0	±0.5
5.	Weight	Kg	2	2.0	±0.5
6.	Weight	Kg	3	3.0	±0.5
7.	Weight	Kg	5	5.1	±0.5

**CALIBRATION RESULT:**

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: 

Date: 11.02.2021

**SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First floor , Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087.