

GSTIN : 29ATZPG5568K1ZL



Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR **REMI** GROUP REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS 24x7

Simply - 27 91148004037

CALIBRATION CERTIFICATE

Ref.

Date :

CUSTOMER: **IQURE BIOMEDICAL SYSTEMS Pvt Ltd**
13/1 2nd Floor 1st Main Road
KHM Block Ganganagar
R T Nagar Post Bengaluru – 560032, Karnataka

Certificate No. : 97525/A

Date: 12/02/2021

Description Of The Item: **Refrigerator**

Model / Make: Whirlpool /174

Sl. No.: 4021/w1850/21402

Range: +2 °C to +8 °C

Details Of Equipment Used For Calibration

Sl. No	Description	Certified By	Certificate No.	Validity
1	Digital Temp. Indicator Range – 100°C to +300°C	STQC Calibration Lab ETDC Bangalore , Govt of India	CR/RF/51147	14/09/2021
2	Glass Thermometer	RRSL STANDARS LABORATORY	RRSL/B/366/O/ 408/2020-21	07/09/2021

Observation

Sl. No.	Instrument Temp. Displayed (° C)	Indicator Temp. (°C)	Temperature Deviation (° C)	Remarks
1	+ 2° C	+ 3° C	01	OK
2	+ 4° C	+ 3° C	01	OK
3	+ 8° C	+ 7° C	01	OK

Next Calibration

Due Date: 11/08/2021

- Remarks:
1. Calibration traceable to National Standards.
 2. Overall Deviation is within specified accuracy hence, instrument is OK.
 3. Results are related only to the item calibrated.
 4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G

GSTIN : 29ATZPG5568K1ZL



Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR REMI GROUP REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS 24x7

Simply - 27 94486003

CALIBRATION CERTIFICATE

Ref.

Date :

CUSTOMER: IQURE BIOMEDICAL SYSTEMS Pvt Ltd

Certificate No. : 97526/A

13/1 2nd Floor 1st Main Road

Date: 12/02/2021

KHM Block Ganganagar

R T Nagar Post Bengaluru – 560032, Karnataka

Description Of The Item: **Refrigerator**

Model / Make: Whirlpool /220

Sl. No.: /INA993103108

Range: +2 °C to +8 °C

Details Of Equipment Used For Calibration

Sl. No	Description	Certified By	Certificate No.	Validity
1	Digital Temp. Indicator Range – 100°C to +300°C	STQC Calibration Lab ETDC Bangalore , Govt of India	CR/RF/51147	14/09/2021
2	Glass Thermometer	RRSL STANDARS LABORATORY	RRSL/B/366/O/ 408/2020-21	07/09/2021

Observation

Sl. No.	Instrument Temp. Displayed (° C)	Indicator Temp. (°C)	Temperature Deviation (° C)	Remarks
1	+ 2° C	+ 3° C	01	OK
2	+ 4° C	+ 5° C	01	OK
3	+ 8° C	+ 7.5° C	0.5	OK

Next Calibration

Due Date: 11/08/2021

Remarks: 1. Calibration traceable to National Standards.

2. Overall Deviation is within specified accuracy hence, instrument is OK.

3. Results are related only to the item calibrated.

4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G

GSTIN : 29ATZPG5568K1ZL



Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR **REMI GROUP** REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS **24x7**

Simply - 27

CALIBRATION CERTIFICATE

Ref.

Date :

CUSTOMER: **IQURE BIOMEDICAL SYSTEMS Pvt Ltd**

Certificate No. : 97527/A

13/1 2nd Floor 1st Main Road

Date: 12/02/2021

KHM Block Ganganagar

R T Nagar Post Bengaluru – 560032, Karnataka

Description Of The Item: **Refrigerator**

Model / Make: **Whirlpool /220**

Sl. No.: **I11A00206D19A0615**

Range: **+2°C to +8°C**

Details Of Equipment Used For Calibration

Sl. No	Description	Certified By	Certificate No.	Validity
1	Digital Temp. Indicator Range – 100°C to +300°C	STQC Calibration Lab ETDC Bangalore , Govt of India	CR/RF/51147	14/09/2021
2	Glass Thermometer	RRSL STANDARS LABORATORY	RRSL/B/366/O/ 408/2020-21	07/09/2021

Observation

Sl. No.	Instrument Temp. Displayed (° C)	Indicator Temp. (°C)	Temperature Deviation (° C)	Remarks
1	+ 2° C	+ 3° C	01	OK
2	+ 4° C	+ 4.5° C	0.5	OK
3	+ 8° C	+ 8.5° C	0.5	OK

Next Calibration

Due Date: **11/08/2021**

Remarks: 1. Calibration traceable to National Standards.

2. Overall Deviation is within specified accuracy hence, instrument is OK.

3. Results are related only to the item calibrated.

4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G

GSTIN : 29ATZPG5568K1ZL



Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR **REMI** GROUP REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS **24x7**

Simply - 27

CALIBRATION CERTIFICATE

Ref.

Date :

CUSTOMER: **IQURE BIOMEDICAL SYSTEMS Pvt Ltd**

Certificate No. : 97528/A

13/1 2nd Floor 1st Main Road

Date: 12/02/2021

KHM Block Ganganagar

R T Nagar Post Bengaluru – 560032, Karnataka

Description Of The Item: **Digital Thermometer**

Model / Make: **PM-10**

Sl. No.: **1 01**

Range: **-50°C to +70°C**

Details Of Equipment Used For Calibration

Sl. No	Description	Certified By	Certificate No.	Validity
1	Digital Temp. Indicator Range – 100°C to +300°C	STQC Calibration Lab ETDC Bangalore , Govt of India	CR/RF/51147	14/09/2021
2	Glass Thermometer	RRSL STANDARS LABORATORY	RRSL/B/366/O/ 408/2020-21	07/09/2021

Observation

Sl. No	Instrument Temp. Displayed (Deg. Cel.)	Indicator Temp. (Deg. Cel.)	Temperature Deviation	Remarks
1	+ 2° C	+ 3° C	01	OK
2	+ 4° C	+ 3° C	01	OK
3	-20° C	- 21° C	01	OK

Next Calibration

Due Date: **11/08/2021**

Remarks: 1. Calibration traceable to National Standards.

2. Overall Deviation is within specified accuracy hence, instrument is OK.

3. Results are related only to the item calibrated.

4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G



CALIBRATION CERTIFICATE

Ref.

Date :

CUSTOMER: **IQUIRE BIOMEDICAL SYSTEMS Pvt Ltd**
13/1 2nd Floor 1st Main Road
KHM Block Ganganagar
R T Nagar Post Bengaluru – 560032, Karnataka

Certificate No. : 97529/A

Date: 12/02/2021

Description Of The Item: **Digital Thermometer**

Model / Make: **PM-10**

Sl. No.: 1 02

Range: **-50°C to +70°C**

Details Of Equipment Used For Calibration

Sl. No	Description	Certified By	Certificate No.	Validity
1	Digital Temp. Indicator Range – 100°C to +300°C	STQC Calibration Lab ETDC Bangalore , Govt of India	CR/RF/51147	14/09/2021
2	Glass Thermometer	RRSL STANDARS LABORATORY	RRSL/B/366/O/ 408/2020-21	07/09/2021

Observation

Sl. No	Instrument Temp. Displayed (Deg. Cel.)	Indicator Temp. (Deg. Cel.)	Temperature Deviation	Remarks
1	+ 2° C	+ 3° C	01	OK
2	+ 4° C	+ 5° C	01	OK
3	-20° C	- 19° C	01	OK

Next Calibration

Due Date: **11/08/2021**

Remarks: 1. Calibration traceable to National Standards.

2. Overall Deviation is within specified accuracy hence, instrument is OK.

3. Results are related only to the item calibrated.

4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G

GSTIN : 29ATZPG5568K1ZL



Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR **REMI** GROUP REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS 24x7

Simply - 27 9448004037

CALIBRATION CERTIFICATE

Ref.

Date :

CUSTOMER: **IQURE BIOMEDICAL SYSTEMS Pvt Ltd**

Certificate No. : 97530/A

13/1 2nd Floor 1st Main Road

Date: 12/02/2021

KHM Block Ganganagar

R T Nagar Post Bengaluru – 560032, Karnataka

Description Of The Item: **Digital Thermometer**

Model / Make: PM-10

Sl. No.: 1 03

Range: -50°C to +70°C

Details Of Equipment Used For Calibration

Sl. No	Description	Certified By	Certificate No.	Validity
1	Digital Temp. Indicator Range – 100°C to +300°C	STQC Calibration Lab ETDC Bangalore , Govt of India	CR/RF/51147	14/09/2021
2	Glass Thermometer	RRSL STANDARS LABORATORY	RRSL/B/366/O/ 408/2020-21	07/09/2021

Observation

Sl. No	Instrument Temp. Displayed (Deg. Cel.)	Indicator Temp. (Deg. Cel.)	Temperature Deviation	Remarks
1	+ 2° C	+ 3° C	01	OK
2	+ 4° C	+ 4.5° C	0.5	OK
3	-20° C	- 18° C	02	OK

Next Calibration

Due Date: 11/08/2021

Remarks: 1. Calibration traceable to National Standards.

2. Overall Deviation is within specified accuracy hence, instrument is OK.

3. Results are related only to the item calibrated.

4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G

GSTIN : 29ATZPG5568K1ZL



Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR **REMI GROUP** REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS 24x7

Simply - 27 9448004037

CALIBRATION CERTIFICATE

Ref.

Date :

CUSTOMER: **IQURE BIOMEDICAL SYSTEMS Pvt Ltd**

Certificate No. : 97531/A

13/1 2nd Floor 1st Main Road

Date: 12/02/2021

KHM Block Ganganagar

R T Nagar Post Bengaluru – 560032, Karnataka

Description of the Item: **Centrifuge**

Model/Make: **REMI -R 8 C Plus**

SI. No.: ZEBN-04645

Range: 6000 RPM

Details of Equipment Used For Calibration

SI No.	Description	Certified By	Certificate No.	Validity
1	Digital Tacho Meter	Adersha Calibration Services Ba'lore	ACS/MC/20-21/102551	04/08/2021

Observation

SI. No	Instrument speed . Displayed (RPM)	Speed Reading on Digital Tachometer (RPM)	Difference in Reading (RPM)	Remarks
1	1000	1012	12	OK
2	2500	2520	20	OK
3	3500	3530	30	OK
4	4500	4543	43	OK
5	5500	5554	54	OK

Next Calibration

Due Date: 11/08/2021

Remarks: 1. Calibration traceable to National Standards.

2. Overall Deviation is within specified accuracy hence, instrument is OK.

3. Results are related only to the item calibrated.

4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G

GSTIN : 29ATZPG5568K1ZL



Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR REMI GROUP REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS 24x7 CALIBRATION CERTIFICATE

Simply - 27

CUSTOMER: **IQURE BIOMEDICAL SYSTEMS Pvt Ltd**
13/1 2nd Floor 1st Main Road
KHM Block Ganganagar
R T Nagar Post Bengaluru – 560032, Karnataka

Certificate No. : 97532/A...
Date: 12/02/2021

Description of The Item: **PIPPTTES**
Make **MENLETTE**

Range --- 5-50 ul

SL NO- MD00417/19

Details Of Equipment Used For Calibration

Sl. No	Description	Certified By	Certificate No.	Validity
1	Brass Weight box 1 mg to 200+ 200 gms, 500g +1 kg +5 kg	RRSL(Weights & Measures) Bangalore Govt of India	RRSL/B/364/O/406/2020-21	07/09/2021
2	Electronic Weighing Machine	RRSL(Weights & Measures) Bangalore Govt of India	RRSL/B/369/O/411/2020-21	07/09/2021

Observation

Sl. No.	Set Weight (g)	Measured Weight (mili g)	Difference in (mili g)	Remarks
1	5	10.0	0.0	OK
2	5	20.0	00	OK
3	5	50.1	0.1	OK

The Equipment is Working Satisfactorily.

Next Calibration Due Date: 11/08/2021

Remarks: 1. Calibration traceable to National Standards.

2. Overall Deviation is within specified accuracy hence, instrument is OK.

3. Results are related only to the item calibrated.

4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G

GSTIN : 29ATZPG5568K1ZL



Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH

Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR REMI GROUP REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS

CALIBRATION CERTIFICATE

Simply - 27 9668004037

/// ///

CUSTOMER: IQURE BIOMEDICAL SYSTEMS Pvt Ltd

13/1 2nd Floor 1st Main Road

KHM Block Ganganagar

R T Nagar Post Bengaluru – 560032, Karnataka

Certificate No. : 97533/A

Date: 12/02/2021

Description of The Item: PIPPTTES

Range --- 100-1000 ul

Make DRAGON-LAB

SL NO- YE202AM0109301

Details Of Equipment Used For Calibration

Sl. No	Description	Certified By	Certificate No.	Validity
1	Brass Weight box 1 mg to 200+ 200 gms, 500g +1 kg +5 kg	RRSL(Weights & Measures) Bangalore Govt of India	RRSL/B/364/O/406/2020-21	07/09/2021
2	Electronic Weighing Machine	RRSL(Weights & Measures) Bangalore Govt of India	RRSL/B/369/O/411/2020-21	07/09/2021

Observation

Sl. No.	Set Weight (g)	Measured Weight (mili g)	Difference in (mili g)	Remarks
1	100	100.0	0.0	OK
2	500	500.1	0.1	OK
3	1000	1000.1	0.1	OK

The Equipment is Working Satisfactorily.

Next Calibration Due Date: 11/08/2021

Remarks: 1. Calibration traceable to National Standards.

2. Overall Deviation is within specified accuracy hence, instrument is OK.

3. Results are related only to the item calibrated.

4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G

GSTIN : 29ATZPG5568K1ZL



Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR REMI GROUP REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS CALIBRATION CERTIFICATE

Simply - 27 9448006037

CUSTOMER: IQURE BIOMEDICAL SYSTEMS Pvt Ltd
13/1 2nd Floor 1st Main Road
KHM Block Ganganagar
R T Nagar Post Bengaluru – 560032, Karnataka

Certificate No. : 97534/A
Date: 12/02/2021

Description of The Item: PIPPTTES
Make MENLETTE -MERIL

SL NO- MD50707/19

Range --- 100-1000 ul

Details Of Equipment Used For Calibration

Sl. No	Description	Certified By	Certificate No.	Validity
1	Brass Weight box 1 mg to 200+ 200 gms, 500g +1 kg +5 kg	RRSL(Weights & Measures) Bangalore Govt of India	RRSL/B/364/O/406/2020-21	07/09/2021
2	Electronic Weighing Machine	RRSL(Weights & Measures) Bangalore Govt of India	RRSL/B/369/O/411/2020-21	07/09/2021

Observation

Sl. No.	Set Weight (g)	Measured Weight (mili g)	Difference in (mili g)	Remarks
1	100	100.1	0.1	OK
2	500	500.0	0.0	OK
3	1000	1000.1	0.1	OK

The Equipment is Working Satisfactorily.

Next Calibration Due Date: 11/08/2021

- Remarks: 1. Calibration traceable to National Standards.
2. Overall Deviation is within specified accuracy hence, instrument is OK.
3. Results are related only to the item calibrated.
4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G