

GSTIN : 29ATZPG5568K1ZL



# Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH  
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR **REMI GROUP** REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS 24x7

Simply - 27

9448504037

## CALIBRATION CERTIFICATE

Date : .....

Certificate No. : 97531/A

Date: 12/02/2021

Ref.

CUSTOMER: **IQURE BIOMEDICAL SYSTEMS Pvt Ltd**

# 13/1 2<sup>nd</sup> Floor 1<sup>st</sup> Main Road

KHM Block Ganganagar

R T Nagar Post Bengaluru – 560032, Karnataka

Description of the Item: **Centrifuge**

Model/Make: **REMI -R 8 C Plus**

Sl. No.: ZEBN-04645

Range: 6000 RPM

### Details of Equipment Used For Calibration

Sl No.	Description	Certified By	Certificate No.	Validity
1	Digital Tacho Meter	Adersha Calibration Services Ba'lore	ACS/MC/20-21/102551	04/08/2021

### Observation

Sl. No	Instrument speed . Displayed (RPM )	Speed Reading on Digital Tachometer (RPM)	Difference in Reading (RPM)	Remarks
1	1000	1012	12	OK
2	2500	2520	20	OK
3	3500	3530	30	OK
4	4500	4543	43	OK
5	5500	5554	54	OK

Next Calibration

Due Date: 11/08/2021

Remarks: 1. Calibration traceable to National Standards.

2. Overall Deviation is within specified accuracy hence, instrument is OK.

3. Results are related only to the item calibrated.

4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G