

GSTIN : 29ATZPG5568K1ZL



# Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH  
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR REMI GROUP REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS 24x7

Simply - 27 94483003

## CALIBRATION CERTIFICATE

Ref.

Date : .....

CUSTOMER: **IQURE BIOMEDICAL SYSTEMS Pvt Ltd**  
# 13/1 2<sup>nd</sup> Floor 1<sup>st</sup> Main Road  
KHM Block Ganganagar  
R T Nagar Post Bengaluru – 560032, Karnataka

Certificate No. : 97526/A  
Date: 12/02/2021

Description Of The Item: **Refrigerator**

Model / Make: Whirlpool /220

Sl. No.: /NA993103108

Range: +2 °C to +8 °C

### Details Of Equipment Used For Calibration

Sl. No	Description	Certified By	Certificate No.	Validity
1	Digital Temp. Indicator Range – 100°C to +300°C	STQC Calibration Lab ETDC Bangalore , Govt of India	CR/RF/51147	14/09/2021
2	Glass Thermometer	RRSL STANDARS LABORATORY	RRSL/B/366/O/ 408/2020-21	07/09/2021

### Observation

Sl. No.	Instrument Temp. Displayed (° C)	Indicator Temp. (°C)	Temperature Deviation (° C)	Remarks
1	+ 2° C	+ 3° C	01	OK
2	+ 4° C	+ 5° C	01	OK
3	+ 8° C	+ 7.5° C	0.5	OK

Next Calibration Due Date: 11/08/2021

- Remarks:
1. Calibration traceable to National Standards.
  2. Overall Deviation is within specified accuracy hence, instrument is OK.
  3. Results are related only to the item calibrated.
  4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G