

GSTIN : 29ATZPG5568K1ZL



Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR REMI GROUP REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS 24x7

Simply - 27

CALIBRATION CERTIFICATE

Ref.

Date :

CUSTOMER: **IQUIRE BIOMEDICAL SYSTEMS Pvt Ltd**
13/1 2nd Floor 1st Main Road
KHM Block Ganganagar
R T Nagar Post Bengaluru – 560032, Karnataka

Certificate No. : 97529/A

Date: 12/02/2021

Description Of The Item: **Digital Thermometer**

Model / Make: **PM-10**

Sl. No.: 1 02

Range: -50°C to +70°C

Details Of Equipment Used For Calibration

Sl. No	Description	Certified By	Certificate No.	Validity
1	Digital Temp. Indicator Range – 100°C to +300°C	STQC Calibration Lab ETDC Bangalore , Govt of India	CR/RF/51147	14/09/2021
2	Glass Thermometer	RRSL STANDARS LABORATORY	RRSL/B/366/O/ 408/2020-21	07/09/2021

Observation

Sl. No	Instrument Temp. Displayed (Deg. Cel.)	Indicator Temp. (Deg. Cel.)	Temperature Deviation	Remarks
1	+ 2° C	+ 3° C	01	OK
2	+ 4° C	+ 5° C	01	OK
3	-20° C	- 19° C	01	OK

Next Calibration

Due Date: 11/08/2021

Remarks: 1. Calibration traceable to National Standards.

2. Overall Deviation is within specified accuracy hence, instrument is OK.

3. Results are related only to the item calibrated.

4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G