

GSTIN : 29ATZPG5568K1ZL



Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR REMI GROUP REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS ^{24x7} CALIBRATION CERTIFICATE

Simply - 27 9668004037

CUSTOMER: **IQURE BIOMEDICAL SYSTEMS Pvt Ltd**
13/1 2nd Floor 1st Main Road
KHM Block Ganganagar
R T Nagar Post Bengaluru – 560032, Karnataka

Certificate No. : 97533/A
Date: 12/02/2021

Description of The Item: **PIPPTTES**
Make **DRAGON-LAB**

Range --- 100-1000 ul

SL NO- YE202AM0109301

Details Of Equipment Used For Calibration

| Sl. No | Description | Certified By | Certificate No. | Validity |
|--------|---|---|--------------------------|------------|
| 1 | Brass Weight box 1 mg to 200+ 200 gms, 500g +1 kg +5 kg | RRSL(Weights & Measures) Bangalore Govt of India | RRSL/B/364/O/406/2020-21 | 07/09/2021 |
| 2 | Electronic Weighing Machine | RRSL(Weights & Measures) Bangalore Govt of India | RRSL/B/369/O/411/2020-21 | 07/09/2021 |

Observation

| Sl. No. | Set Weight (g) | Measured Weight (mili g) | Difference in (mili g) | Remarks |
|---------|----------------|--------------------------|------------------------|---------|
| 1 | 100 | 100.0 | 0.0 | OK |
| 2 | 500 | 500.1 | 0.1 | OK |
| 3 | 1000 | 1000.1 | 0.1 | OK |

The Equipment is Working Satisfactorily.

Next Calibration Due Date: 11/08/2021

Remarks: 1. Calibration traceable to National Standards.

2. Overall Deviation is within specified accuracy hence, instrument is OK.

3. Results are related only to the item calibrated.

4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G