

GSTIN : 29ATZPG5568K1ZL



Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR REMI GROUP REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS 24x7 CALIBRATION CERTIFICATE

Simply - 27 944800407

CUSTOMER: **IQURE BIOMEDICAL SYSTEMS Pvt Ltd**
13/1 2nd Floor 1st Main Road
KHM Block Ganganagar
R T Nagar Post Bengaluru – 560032, Karnataka

Certificate No. : 97534/A.....
Date: 12/02/2021

Description of The Item: **PIPPTES**
Make **MENLETTE -MERIL**

Range --- 100-1000 ul

SL NO- MD50707/19

Details Of Equipment Used For Calibration

Sl. No	Description	Certified By	Certificate No.	Validity
1	Brass Weight box 1 mg to 200+ 200 gms, 500g +1 kg +5 kg	RRSL(Weights & Measures) Bangalore Govt of India	RRSL/B/364/O/406/2020-21	07/09/2021
2	Electronic Weighing Machine	RRSL(Weights & Measures) Bangalore Govt of India	RRSL/B/369/O/411/2020-21	07/09/2021

Observation

Sl. No.	Set Weight (g)	Measured Weight (mili g)	Difference in (mili g)	Remarks
1	100	100.1	0.1	OK
2	500	500.0	0.0	OK
3	1000	1000.1	0.1	OK

The Equipment is Working Satisfactorily.

Next Calibration Due Date: **11/08/2021**

- Remarks:
1. Calibration traceable to National Standards.
 2. Overall Deviation is within specified accuracy hence, instrument is OK.
 3. Results are related only to the item calibrated.
 4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G