

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/ULR-CC2194200000000510F		DISCIPLINE	THERMAL	
Calibration Date	30.09.2020			
Calibration Due	29.09.2021			
Customer Name & Address				
SREE METRO DIAGNOSTIC CENTRE 16/41, KRISHNA ROAD, NEW PERUNGALATHUR, CHENNAI-63				
Details of Device Under Calibration (DUC)				
Description	: REFRIGERATOR	Make & Model	: SAMSUNG & NA	
Range	: 2°C to 8°C	Sr. No	: NA	
Least Count	: 0.1°C	Barcode No	: SMDC/LAB-3/REF/02	
DUC Condition	: Satisfactory	Location	: LAB-3	
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature: 25° C	Relative Humidity	51% Rh	
Calibration Procedure No	CP-T02-2018	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	4 Wire RTD Sensor	Tempsen Devices / TD27405	M.K.S/TH/18-19/51-02	06.07.2021
2	Digital mutimeter	GW Instek / GES 817082	HT/1732/001	06.07.2021

TEST REPORT: SBS/TR/ULR CC2194200000000510F

S.No.	STANDARD INSTRUMENTS READINGS	DEVICE UNDER CALIBRATION READINGS	DEVIATION	EXPANDED UNCERTAINTY (±)
	°C	°C	°C	°C
1	2.00	1.9	-0.1	0.33
2	4.00	3.9	-0.1	0.33
3	5.00	5.1	0.1	0.34
4	7.00	7.2	0.2	0.33
5	8.00	8.2	0.2	0.33

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
 2. The user should determine the suitability of the instrument for its intended use.
 3. The recalibration interval should be determined on the user requirement.
 4. The results stated in this certificate relate only to the item calibrated.
 5. Any correction / Modification in this calibration report kindly inform with in 30 days.
- factor k=2.00 .

Calibrated By
(Signature)
(Calibration Engineer)



Authorised Signatory
(Signature)
 (Technical Manager) (Chief Executive)



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/ULR-CC219420000000509F DISCIPLINE THERMAL

Calibration Date 30.09.2020
Calibration Due 29.09.2021

Customer Name & Address
SREE METRO DIAGNOSTIC CENTRE
16/41, KRISHNA ROAD, NEW PERUNGALATHUR, CHENNAI-63

Details of Device Under Calibration (DUC)

Description : REFRIGERATOR Make & Model : GODREJ & NA
Range : 2°C to 8°C Sr. No : NA
Least Count : 0.1°C Barcode No : SMDC/LAB-2/REF/01
DUC Condition : Satisfactory Location : BLOOD BANK

Environmental Conditions & Calibration Procedure Details

Environmental Details Temperature: 25° C Relative Humidity 51% Rh
Calibration Procedure No CP-T02-2018 Calibration done at ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	4 Wire RTD Sensor	Tempsen Devices / TD27405	M.K.S/TH/18-19/51-02	06.07.2021
2	Digital mutimeter	GW Instek / GES 817082	HT/1732/001	06.07.2021

TEST REPORT: SBS/TR/ULR CC219420000000509F

S.No.	STANDARD INSTRUMENTS READINGS	DEVICE UNDER CALIBRATION READINGS	DEVIATION	EXPANDED UNCERTAINTY (±)
	°C	°C	°C	°C
1	2.00	2.0	0.0	0.33
2	4.00	4.0	0.0	0.33
3	5.00	5.1	0.1	0.34
4	7.00	7.3	0.3	0.33
5	8.00	8.3	0.3	0.35

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
 - 2.The user should determine the suitability of the instrument for its intended use.
 - 3.The recalibration interval should be determined on the user requirement.
 - 4.The results stated in this certificate relate only to the item calibrated.
 - 5.Any correction / Modification in this calibration report kindly inform with in 30 days.
- factor k=2.00 .

Calibrated By

(Calibration Engineer)



Authorised Signatory

(Technical Manager)

(Chief Executive)

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First floor , Poomagal Main Road, Ekkattulhangal, Chennai - 600 032, Tel: 044 - 2225 2087.