

	WBC	RBC	HGB	HCT	PLT	MPV
	10 ³ /mm ³	10 ⁶ /mm ³	g/dL	%	10 ³ /mm ³	μm ³

Average	9.3	4.48	12.6	39.6	235	8.8
%BAL	4.0	1.04	9.2	1.5	0	0.9
CV	0.8	2.43	0.3	2.6	5	2.2
New Coef	0.97	0.84	1.19	1.02	1.15	0.95
Old Coef	1.01	0.83	1.08	1.04	1.15	0.96
Target	8.9	4.53	13.9	39.0	235	8.7
A 1	8.9	3.81	12.0	33.4	202	8.7
2	9.4	4.61	12.7	40.7	252	8.7
3	9.2	4.54	12.6	40.2	238	8.8
4	9.3	4.47	12.6	39.4	230	9.1
5	9.2	4.32	12.6	38.0	220	8.6
6						
7						
8						
9						
10						
11						

LOT

CX 454

Rev 1

CAL

 (Exp.) 2021-02-05
(YYY - MM - DD)

Minilysebio

PARAMETRES PARAMETERS	UNITES UNITS	MICROS						PENTRA			TOLERANCES TOLERANCE
		ES60	CRP 200			60 8P 60 18P					
GB WBC	10 ³ /mm ³ , 10 ⁹ /l	8.9	8.8			9.0					± 0.2
GR RBC	10 ⁶ /mm ³ , 10 ¹² /l	4.53	4.52			4.55					± 0.06
HB HGB	g/dl	13.9	13.4			13.8					± 0.2
		139	134			138					± 2
	mmol/l	8.63	8.34			8.57					± 0.12
HT HCT	%	39.0	38.9			39.6					± 1.0
PLAQ. PLTS	10 ³ /mm ³ , 10 ⁹ /l	0.390	0.389			0.396					± 0.010
VPM MPV	µm ³ , fl	235	242			244					± 10
		8.7	8.2			9.2					± 0.5

ABX Minocal



CAL

(Exp.) 2021-02-05
 (MM - MM - DD) YYY

LOT

CX 454
 Rev 1

PARAMETRES	UNITS	MICROS		MICROS		TOLERANCES
		Soll Vet ABC	ESV ABC Plus	ESV 60		
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	8.7	8.7			± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.54	4.52			± 0.06
HB HGB	g/dl	13.6	13.5			± 0.2
	g/l	136	135			± 2
HT HCT	mml/l	8.45	8.38			± 0.12
	%	39.0	38.9			± 1.0
PLAQ. PLTS	l/l	0.390	0.389			± 0.010
	10 ² /mm ³ ;10 ⁹ /l	240	249			± 10
VPM MPV	µm ³ fl	N/A	9.0			± 0.5

Ref: TEMP-0387 Rev.46 FRONT / RECTO 1300060881