

Registered Office:
BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Irungattukottai,
Thandalam Village, Sriperumbudur - 602105.
Ph: +91 - 44 - 3034 1911 / 12 / 13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN - U29297TN2003PTC051492



Date: 30/10/2020

CALIBRATION CERTIFICATE

This is to certify that BTS - 350 Photometer bearing S.No:8017512371 installed at MOLXPRT DIAGNOSTICS PVT. LTD., GREATER NOIDA has been calibrated on 30/10/2020

Calibration Values:

LIGHT INTENSITY TEST



LED	TINT	MAIN	REFERENCE
340	17	980008	282030
405	13	982127	738139
505	28	976749	758624
535	12	950054	522244
560	12	939964	586812
600	22	965458	608412
635	13	980738	629844
670	13	945000	694697

PERISTALTIC PUMP ADJUSTMENT TEST

VALUE	CURRENT	NEW
SAMPLE VOLUME	45877	45877
POSITIONING	2378	2378
PUMP DELAY	2	2

Next calibration is due on 29/09/2021

For BIOSYSTEMS DIAGNOSTICS Pvt. Ltd.,



AUTHORISED SIGNATORY

Corporate Office:
BioSystems S.A. Costa Brava 30, 08030 Barcelona (Spain)
Tel +34-93 311 00 00, Fax+34-93 346 77 99
<http://www.biosystems.es>

ISO 13485