

Parashar Micro Measurement Pvt. Ltd.



The way to Technology
(18 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info

Calibration Certificate

Page 1 of 1

Certificate Number : PMM/CF/675773		Date of Request	28.01.2021
Calibrated For	M/s. Molxper Diagnostics Pvt. Ltd.	Calibrated on	28.01.2021
	106, Frist Floor Mahagun Mywoods Mart Opp. Gaur City 2, Plot GH04, Sector 16C Greater Noida. West.	Valid Upto	27.01.2022
		Cert. Issue Date	06.02.2021

Description of Equipments

Nomenclature	Centrifuge	Accuracy/Class	Not Mentioned
Range	0 - 3500 rpm	Condition of UUC	Physically Ok
Resolution / Least Count	-----	Serial No.	ZEIN - 30554
Make & Model No.	Remi	Party ID Mark No.	-----
Material / Type	-----	Location	Lab.

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Lutron	PMM/TM/01	PMMPL, Noida	PMM/TM/1201	07.09.2021

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

Calibration Performed at	Site	Environmental	Temperature = $23 \pm 5^\circ \text{C}$
Calibration Procedure No.	CP/MMV/S/03	Condition(s)	Humidity = 30 to 75 % RH
Reference Standard / Guideline	-----		

Mechanical Calibration

CALIBRATION RESULTS

Sr.No.	Step	Actual Observed Value in Std. (in rpm)	\pm Uncertainty in Measurement (in rpm)
1	1	1529	45
2	2	2134	45
3	3	2627	45
4	4	2943	45
5	5	3486	45


UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$)

Tapan

Calibrated By : Tapan Kumar
Designation : Calibration Engineer



Moh

Approved By : Moh. Iqbal
Designation : Tech. Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.