

# Parashar Micro Measurement Pvt. Ltd.

B-59, SECTOR-64, NOIDA (U.P.) 201 307  
 PH: 0120-4252360-65, Email: pmmplindia@gmail.com  
 Website: www.pmmpl.in, www.parashar.info  
 (NABL ACCREDITED LABORATORY)



The Best in Technology  
 18 Years of Excellence



## Calibration Certificate

ULR - CC279521000002827F

Certificate Number - PMM/MP/11303/1		Page 1 of 1	
Calibrated For	M/s. Molixpert Diagnostics Pvt Ltd	UUC Received on	03.02.2021
	106 First Floor Mahagun Mywoods Mart Opp Gaur City 2 Plot GH04, Sector 16C Greater Noida West	Calibrated on	03.02.2021
		Valid Upto	02.08.2021
		Cert. Issue Date	08.02.2021

### Description of Equipments

Nomenclature	Micro Pipette	Accuracy/Class	Not Mentioned
Range	10 - 100 µl	Condition of UUC	Physically Ok
Resolution / Least Count	1 µl	Serial No.	YE186A20015829
Make & Model No	Dragon Lab	Party ID Mark No.	-----
Material / Type	-----	Location	-----

### Master Equipment / Standard Used

Sr.No	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Weighing Balance	Sartorius	PMM/WM/06	PMMPL, Noida	PMM/WB/1181	19.05.2021

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

Calibration Performed at	Laboratory	Environmental Condition(s)	Temperature = 23 ± 2° C
Calibration Procedure No.	CP/M/MV/04		Humidity = 30 to 75 % RH
Reference Standard / Guideline	ISO-8655-6		

### Mechanical Calibration


### CALIBRATION RESULTS


Sr.No.	Nominal Value in UUC (in µl)	Actual Observed Value Mean Value in Std. (in µl)	Error (in µl)	Mean Average Value at 27°C (in µl)
1	10	9.97	0.03	9.97
2	50	49.87	0.13	49.88
3	100	99.74	0.26	99.76


UUC - Unit Under Calibration.

Std - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor  $k = 2$ ) = ± 0.60 µl

Calibrated By:   
 Designation: Calibration Engineer



Approved By:   
 Designation: Tech. Manager / Auth. Sign.

- Conditions:
1. This certificate refers only to the particular item submitted for calibration.
  2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
  3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.