

Parashar Micro Measurement Pvt. Ltd.

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info
(NABL ACCREDITED LABORATORY)



The way to Technology
(18 Years of Excellence)



CC - 2795

Calibration Certificate

ULR - CC279521000002828F		Page 1 of 1	
Certificate Number : PMM/MP/11303/2		UUC Received on	03.02.2021
Calibrated For	M/s. Molxper Diagnostics Pvt. Ltd. 106, First Floor Mahagun Mywoods Mart Opp. Gaur City 2, Plot GH04, Sector 16C Greater Noida. West.	Calibrated on	03.02.2021
		Valid Upto	02.08.2021
		Cert. Issue Date	08.02.2021

Description of Equipments

Nomenclature	Micro Pipette	Accuracy/Class	Not Mentioned
Range	100 - 1000 µl	Condition of UUC	Physically Ok
Resolution / Least Count	1 µl	Serial No.	WI473590
Make & Model No.	P' Fact	Party ID Mark No.	-----
Material / Type	-----	Location	-----

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Weighing Balance	Sartorius	PMM/WM/06	PMMPL, Noida	PMM/WB/1181	19.05.2021

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

Calibration Performed at	Laboratory	Environmental	Temperature = 23 ± 2° C
Calibration Procedure No.	CP/M/MV/04	Condition(s)	Humidity = 30 to 75 % RH
Reference Standard / Guideline	ISO-8655-6		

Mechanical Calibration

CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in µl)	Actual Observed Value Mean Value in Std. (in µl)	Error (in µl)	Mean Average Value at 27°C (in µl)
1	100	100.30	-0.30	100.32
2	500	499.81	0.19	499.87
3	1000	999.71	0.29	999.85

UUC:- Unit Under Calibration.

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 0.60 µl

Calibrated By : Gordhan Bairwa
Designation : Calibration Engineer



Approved By : Jai Prakash
Designation : Tech. Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.