

MATOSHREE INSTRUMENTS™

Works : EL-107, 'Abhinav', 3rd Floor, T.T.C. Industrial Area, MIDC Mahape, Navi Mumbai - 400710. Mobile No.: 9702578323. Email: matoshree.instruments@gmail.com.

Calibration Certificate

Glass Thermoameter

Certificate No: MI 2012004

08.12.2020 Request Date: **PGM Diagnostic** 08.12.2020 Received on: Shop No. 41/42, G- Wing, Good **Physical Condition:** Service Requested by: Gokul Nagari -II, W.E. Highway, Dattani Road, Kandivali - East, Mumbai - 400101 Details of Instrument/Unit Under Calibration (UUC) 08.12.2020 Calibration Date: Instrument Name: Glass Thermometer 07.12.2021 Calibration Due Date: Manufacturer: Domple Model: Serial No.: Lab Calibrated at: ld. No. **Ambient Conditions** 53.6 %RH ± 5 %RH Relative Humidity Temperature 25.4 ± 2 °C **Calibration Specifications** Calbration Set Point Accuracy Least Count. Range **Parameters** 0, 25, 50, 100 °C ±2°C -10 to 110 °C Temperature Reference Standards Used **Due Date** Cal. Date Certificate No. Make Serial No/Id. No. **Equipment Type** 06.07.2021 07.07.2020 QATM-20070078 MI MI-002 RTD PT-100 Sensor 06.07.2021 07.07.2020 QATM-20070080

Calibration Procedure:

Digital Multimeter

Glass Thermometer is calibrated by comparison with standards.

Observation: Parameter: Temperature

MI-017

Rishabh

Set Point		Standard Readings	Deviation	Uncertainity
Set Point	- ° C	°C	° C	°C
	 	0.2	-0.2	
25	25	25.2	-0.2	
50	49	49.5	-0.5	±0.5
100	99	99.9	-0.9	
100				

Note:

UUC: Unit under calibration

Uncertainty has been calculated for a coverage factor k=2 corresponding to approximately 95% Confidence Level

The Standard maintained are traceable to National/ International Standard through accredited Laboratories. The observations reported represent values at the time of the measurements and under the stated condition. They do not convey any long term stability information.

Authorized By:

Name:

END OF REPORT

Page 01 of 01