

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company


No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCA8053H1ZR • DL No. MDU/4205/208,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE



Customer Name & Address:	M/S. Marson Clinical Lab, 25B, Muhammathiyar Street, Eral - 628 801.	Certificate No:	AMS/LAB/2888807
		Issuing Date:	15/12/2020
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Refrigerator - Reagent	Id. No:	MAR/LAB/REFR/001
Make:	LG	Department:	Laboratory
Model:	GLB 254AM95	Calibrated On:	14/12/2020
Sl. No:	023650	Calibration Due Suggested:	13/12/2021
Type:	Reagent Direct Cool	DUC Condition on Receipt:	OK
Range & Least Count:	0 - 8°C / LC 1°C		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	QC/AMS/DTI/002	22/02/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

Test Results:

Sl.No.	DUC Reading (°C)	Standard Reading (°C)	Correction (°C)
1	2.0	1.30	-0.70
2	-1.0	-0.29	+0.71

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. Srinivas



www.acemedical.in

Approved by:

(Technical Manager)

B. Srinivas

93455 56140, 93455 56148

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCA8053H1ZR

DL No. MDU/4205/20B, MDU/3971/21B

CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Marson Clinical Lab, 25B, Muhammathiyar Street, Eral - 628 801.	Certificate No:	AMS/LAB/2889254
		Issuing Date:	20/02/2021
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Refrigerator - Reagent	Id. No:	MAR/LAB/REFR/001
Make:	LG	Department:	Laboratory
Model:	GLB 254AM95	Calibrated On:	19/02/2021
Sl. No:	023650	Calibration Due Suggested:	18/02/2022
Type:	Reagent Direct Cool	DUC Condition on Receipt:	OK
Range & Least Count:	0 - 8°C / LC 1°C		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	PC - 01 - 21/ATL/0107 - 01	22/01/2022
2	Thermo Hygrometer	HTC	PC - 01 - 21/ATL/0194 - 01	22/01/2022

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.No.	Place	DUC Reading (°C)	Standard Reading (°C)	Correction (°C)
1	Freezer	-5.0	-4.63	+0.37
2		-3.0	-2.78	+0.22
3	Chiller	-1.0	-0.29	+0.71
4		2.0	1.38	-0.62
5	Chamber 1	3.0	2.56	-0.44
6		4.0	3.43	-0.57
7	Chamber 2	4.5	3.98	-0.52
8		5.5	5.06	-0.44

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. S. S.



Approved by:

(Technical Manager)

B. S. S.