

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140



CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. VSV Labs, No.45, Ponmeni Narayanan Street, S.S Colony, Madurai – 625 016	Certificate No:	AMS/LAB/2889274
		Issuing Date:	25/02/2021
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Refrigerator – Reagent	Id. No:	VSV/LAB/RREF/001
Make:	Kenstar	Department:	Laboratory
Model:	Small – Single Door	Calibrated On:	24/02/2021
Sl. No:	----	Calibration Due Suggested:	23/02/2022
Type:	Reagent	DUC Condition on Receipt:	OK
Range & Least Count:	(-)10to8°C / LC 1°C		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	PC/01/21/ATL/0107-01	22/01/2022
2	Thermo Hygrometer	Beetech/088	PC/01/21/ATL/0194-01	22/01/2022

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.No.	Place	DUC Reading (°C)	Standard Reading (°C)	Correction (°C)
1	Freezer	-7.5	-7.09	+0.41
2		-6.5	-6.12	+0.38
3	Chiller	-1.0	-0.79	+0.21
4		1.0	1.24	+0.24
5	Chamber 1	3.0	3.18	+0.18
6		4.0	4.08	+0.08
7	Chamber 2	5.0	5.11	+0.11
8		5.5	5.71	+0.21

Allowable Tolerance: User Defined.

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test
(i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. S. S.



Approved by:

(Technical Manager)

B. S. S.