

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCAS053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140



## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. VSV Labs, No.45, Ponmeni Narayanan Street, S.S Colony, Madurai – 625 016	Certificate No:	AMS/LAB/2888501
		Issuing Date:	26/11/2020
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Micro Pipette	Id. No:	VSV/LAB/FMP/002
Make:	C Poulo	Department:	Laboratory
Model:	Fixed Pipette	Calibrated On:	25/11/2020
Sl. No:	14213986	Calibration Due Suggested:	24/11/2021
Type:	Fixed	DUC Condition on Receipt:	OK
Range & Least Count:	50µl		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	LM-2008863	08/08/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

### Test Results:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	50	50	50	0

Allowable Tolerance: ±1%

Found. Correction of (-) 2.2 µl; Pipette is serviced and Tuned to get the correct volume. Now the correction is 'Zero'  
B. S. S.



Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. S. S.

Approved by:

(Technical Manager)

B. S. S.