

LAB MONTHLY SUMMARY



Lab Name **STL DIAGNOSTIC CENTRE** Lab No **14139**
 Month **January** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 18/01/2021

Date of Report Published : 05/02/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Biosystems	255	119.68	9.01	10.79	153 mg/dl	3.09	1.35
2	UREA II	Urease UV / GLDH II	Biosystems	193	31.36	10.37	3.25	41.8 mg/dl	3.21	0.47
3	CREATININE II	Enzymatic Colorimetric II	Biosystems	21	1.24	15.99	0.20	1.22 mg/dl	-0.10	0.09
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Biosystems	213	2.27	21.23	0.48	1.09 mg/dl	-2.45	0.07
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	209	5.46	10.49	0.57	6.05 g/dl	1.03	0.08
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	204	3.33	9.62	0.32	4.15 g/dl	2.56	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	202	5.61	14.10	0.79	6.99 mg/dl	1.74	0.11
8	CHOLESTROL II	CHOD-PAP II	Biosystems	240	115.43	10.70	12.35	116.1 mg/dl	0.05	1.59
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	208	143.79	12.33	17.73	159 mg/dl	0.86	2.46

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****