



## CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/GHKTA/CEN/H01	Date of Issue	: 06/02/2021
		Calibration Done On	: 04/02/2021
		Calibration Due On	: 03/02/2022

Customer Name	<b>Govt. HOSPITAL, KALPETTA, WAYANAD.</b>
---------------	---

### DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Centrifuge	Kemi / C.4S	27364	-	ICTC Lab

### DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/20-21/102671-01	24/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308-CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

<b>Environmental Conditions</b>	Temperature	27°C	Relative Humidity	55 %
---------------------------------	-------------	------	-------------------	------

### PERFORMANCE DATA

Set Reading	Display Reading	Measured Reading	Tolerance
Minimum	1500 rpm	1501 rpm	± 5 rpm
Medium	2500 rpm	2501 rpm	± 5 rpm
Maximum	4500 rpm	4502 rpm	± 5 rpm

<b>Calibration Status</b>	Pass	✓	Fail	
---------------------------	------	---	------	--

Calibrated By

*Mahesh*  
 Mahesh MC (Service Engineer)

Approved By

*Binoj Thomas*  
 Binoj Thomas (Quality Manager)

