



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/GHKTA/REF/H01	Date of Issue	: 06/02/2021
		Calibration Done On	: 04/02/2021
		Calibration Due On	: 03/02/2022

Customer Name **Govt. HOSPITAL, KALPETTA, WAYANAD.**

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Refrigerator	Samsung / RR19T11CBSE/HL/2020	-	-	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	10055593	ACS/20-21/102669-01	23/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27 °C	Relative Humidity	55 %
---------------------------------	-------------	-------	-------------------	------

PERFORMANCE DATA

Display	Measured	Tolerance
1.3 °C	1.3 °C	± 0.3 °C
2.4 °C	2.3 °C	± 0.3 °C
4.2 °C	4.2 °C	± 0.3 °C
7.2 °C	7.2 °C	± 0.3 °C

Calibration Status	Pass	✓	Fail
---------------------------	------	---	------

Calibrated By

Approved By

Prince CP
 Prince CP (Service Engineer)

Binoj Thomas
 Binoj Thomas (Quality Manager)

