



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/GHKTA/VDRL/H01	Date of Issue	: 06/02/2021
		Calibration Done On	: 04/02/2021
		Calibration Due On	: 03/02/2022

Customer Name	Govt. HOSPITAL, KALPETTA, WAYANAD.
---------------	------------------------------------

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
VDRL Rotator	Rotek Plus / PVD Roid	2014	-	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/20- 21/I02671-01	24/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308CTH	20191602	ACS/MC/20- 21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27°C	Relative Humidity	55 %

PERFORMANCE DATA

Result	Obtained RPM
	223 rpm

Calibration Status	Pass	✓	Fail	
--------------------	------	---	------	--

Calibrated By


 Mahesh MC (Service Engineer)

Approved By


 Binoj Thomas (Quality Manager)

