



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/THQHSY/CEN/D01	Date of Issue	: 06/02/2021
		Calibration Done On	: 04/02/2021
		Calibration Due On	: 03/02/2022
Customer Name	TALUK HEAD QUARTERS HOSPITAL, SULTHAN BATHERY, WAYANAD.		

DETAILS OF DEVICE UNDER CALIBRATION				
Equipment Name	Make / Model	Serial No.	Asset ID	Location
Centrifuge	Remi R – 8C Plus	ZGAN-01713	-	ICTC Lab


DETAILS OF MASTER EQUIPMENT USED					
Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/20-21/I02671-01	24/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308-CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
---------------------------------	-------------	------	-------------------	------

PERFORMANCE DATA			
Set Reading	Display Reading	Measured Reading	Tolerance
Minimum	1501 rpm	1501 rpm	± 5 rpm
Medium	5410 rpm	5411 rpm	± 5 rpm
Maximum	7500 rpm	7502 rpm	± 5 rpm

Calibration Status	Pass	✓	Fail	
---------------------------	------	---	------	--

Calibrated By

 Farzan Sathar (Service Engineer)

Approved By

 Binoj Thomas (Quality Manager)



NB: Results reported are valid at the time of and under the stated conditions of measurements. This report shall not be reproduced without written approval by Medical Engineering & Services.