



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/THQHSY/VDRL/D01	Date of Issue : 06/02/2021
		Calibration Done On : 04/02/2021
		Calibration Due On : 03/02/2022
Customer Name	TALUK HEAD QUARTERS HOSPITAL, SULTHAN BATHERY, WAYANAD.	

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
VDRL Rotator	Rotek Plus / PVD Roid	2001	68050045	Blood Bank ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/20-21/I02671-01	24/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.


Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
--------------------------	-------------	------	-------------------	------

PERFORMANCE DATA


Result	Obtained RPM 229 rpm
--------	-------------------------

Calibration Status	Pass	✓	Fail	
--------------------	------	---	------	--

Calibrated By


 Vivek Mahendran (Service Engineer)

Approved By


 Binoj Thomas (Quality Manager)