



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/THQHSY/REF/D01	Date of Issue : 06/02/2021
		Calibration Done On : 04/02/2021
		Calibration Due On : 03/02/2022

Customer Name	TALUK HEAD QUARTERS HOSPITAL, SULTHAN BATHERY, WAYANAD.
---------------	--

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Refrigerator	Samsung / RR19T11CBSE/HL -01	-	-	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	10055593	ACS/20-21/I02669-01	23/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308CTH	20191602	ACS/MC/20-21/2593- 01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.


Environmental Conditions	Temperature	27 °C	Relative Humidity	55 %
--------------------------	-------------	-------	-------------------	------

PERFORMANCE DATA


Display	Measured	Tolerance
1.3 °C	1.1 °C	± 0.3 °C
2.4 °C	2.4 °C	± 0.3 °C
4.2 °C	4.1 °C	± 0.3 °C
7.2 °C	7.2 °C	± 0.3 °C

Calibration Status	Pass	✓	Fail	
--------------------	------	---	------	--

Calibrated By


 Vivek Mahendran (Service Engineer)



Approved By

 Binoj Thomas (Quality Manager)