


LAB MONTHLY SUMMARY



Lab Name **LIFE CARE DIAGNOSTIC CENTRE** Lab No **5256**
 Month **January** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 24/01/2021

Date of Report Published : 05/02/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Mindray	130	122.96	9.24	11.36	134 mg/dl	0.97	1.99
2	UREA II	Urease UV / GLDH II	Mindray	97	31.55	14.03	4.43	33 mg/dl	0.33	0.90
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Mindray	82	1.30	12.50	0.16	1.4 mg/dl	0.62	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	110	2.03	16.26	0.33	2.35 mg/dl	0.97	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	116	5.47	8.10	0.44	5.67 g/dl	0.45	0.08
6	ALBUMIN II	BCG - Colorimetric II	Mindray	116	3.39	8.94	0.30	2.8 g/dl	-1.95	0.06
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	109	5.47	15.82	0.87	6.2 mg/dl	0.84	0.17
8	CHOLESTROL II	CHOD-PAP II	Mindray	129	115.31	10.01	11.54	111 mg/dl	-0.37	2.03
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	115	144.37	12.47	18.00	171 mg/dl	1.48	3.36

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

******* End of Report *******